The potential for MR-Relaxometry to evaluate absorption of oxygen after per oral administration of oxygenated water

Svein Are Sirirud Vatnehol

Doctoral thesis

Faculty of Medicine

University of Oslo

The Intervention Centre

Oslo University Hospital

and

Oxy Solutions A/S

© Svein Are Sirirud Vatnehol, 2020

*Series of dissertations submitted to the
Faculty of Medicine, University of Oslo*

ISBN 978-82-8377-721-5

All rights reserved. No part of this publication may be
reproduced or transmitted, in any form or by any means, without permission.

Cover: Hanne Baadsgaard Utigard.
Photo cover: Øystein Horgmo / UIO
Print production: Reprosentralen, University of Oslo.

“Journey before destination”

“The Stormlight Archives”

Brandon Sanderson 1975 -

“Dovie’andai se tovyu sagain”

“The Wheel of time”

Robert Jordan 1948 - 2007

Acknowledgements

I would first like to thank the Faculty of Medicine at the University of Oslo and the Intervention Centre for giving me the opportunity to conduct this project, and the Norwegian Research Council for your generous support of my PhD project.

I will forever be in debt to my fantastic team of supervisors, Tryggve Holck Storås, Camilla Haglerød, Per Kristian Hol, Mahmood Amiry-Moghaddam, and Atle Bjørnerud, who have guided me throughout this process. Thank you so much for all your work as you have guided me through the process from being a young aspiring student to having (hopefully) become a full-grown researcher.

In addition to my supervisors, there are several other persons that deserve my deepest thanks for helping me along the way during this journey. My friends and colleagues at the Intervention Centre, and especially my close companion Grethe Løvland, whose fantastic organisational skills have made it possible to scan my volunteers. Also at the Intervention Centre: Hilde Korslund who has contributed with much motivation and good discussions over the years and Jorunn Fraser-Green who provided excellent assistance with oxygen measurements in my final study. I would also like to thank Anne Schad Bergsaker at the University of Oslo, for her assistance with in statistics during the course of my PhD.

Before I even started to think about embarking on a PhD journey, several people had an influence on me that brought me to this stage in my life. Firstly, my parents and family who have given me loving support in all my endeavours, be it dancing, education, or something completely different. My first mentor in - and the man who gave me my love of - MRI, Kjell Inge Gjesdal; I think I learned more about MRI under your tutelage than I ever did as a BSc and MSc student. Today I have the privilege of having several friends that are equally nerdy as me, I believe that meeting you for beer and MRI discussions has helped keep me sane

during the last few years. Also, I would like to thank the researchers and PhD students that have gone before me and allowed me to participate in their studies. Observing your work helped a great deal when it came time for me to start my own project.

Two people that deserve a special thank-you are the late Mr. Robert Jordan, and his “heir” Mr. Brandon Sanderson, whose writing gave me the love of reading, the training in reading and the understanding of language that have made my journey possible. Without your books I would not be in this position today, therefore this dissertation stands as evidence that your writing changes lives.

Finally, and above all, to my wonderful wife Rita, who has stood by my side from before the start of my PhD journey, through all the valleys and hills, and finally to the finished product. And Sverre, the smartest, kindest, and most beautiful boy in the world. It would be impossible to explain how much your support has meant to me. I love you with all of my heart, and this dissertation is for the two of you.

Svein Are Sirirud Vatnehol

Table of Contents

Acknowledgements	II
Table of Contents	IV
Abbreviations and definitions	1
List of Papers.....	3
1. Background.....	4
2. Introduction.....	5
2.1. Usage and transportation of oxygen.....	5
2.1.1. The effect of oxygen on uptake through the gastrointestinal epithelium.....	5
2.1.2. Oxygenated water as an alternative oxygen supply	6
2.2. Aims	11
3. Methods.....	12
3.1. Methodological considerations	12
3.2. Measurement of oxygen concentrations in water by Winkler titration.....	13
3.3. MRI	14
3.3.1. Basic Magnetic Resonance Imaging	14
3.3.2. Spin relaxations, spin-lattice and spin-spin interactions	14
3.3.2.1. T_1 relaxation, spin-lattice relaxation	14
3.3.2.1. T_1 of blood	14
3.3.2.2. T_2^* relaxation.....	15
3.3.2.3. T_2^* of blood	15
3.3.2.4. Evolution of the NMV	16
3.3.3. Quantitative MRI.....	17
3.3.3.1.1. The Look-Locker method.....	19
3.3.3.1.2. Modified Look-Locker Inversion recovery (MOLLI).....	20
3.3.3.2. Effect of technical parameters on T_1 -mapping	22
3.3.3.2.1. Readout duration.....	22
3.3.3.2.2. Flip angle	24

3.3.3.2.3.	Timing method	24
3.3.3.3.	T ₂ * measurements.....	25
3.3.4.	Ethical approval.....	25
3.3.5.	MRI scanner and volunteers.....	25
3.3.6.	MRI scan parameters.....	27
3.3.7.	MRI materials.....	29
3.3.8.	Image analysis	29
3.3.9.	Calculations of limits of detection (LOD).....	30
3.3.10.	Statistical methods.....	32
3.3.10.1.	Paired samples t-test	32
3.3.10.2.	Linear regression analysis.....	32
3.3.10.3.	Bland Altman plots and analysis.....	34
3.3.10.4.	Mixed Linear Model (MLM)	36
4.	Results.....	37
5.	Summary of Papers	41
6.	Discussion	45
6.1.	Effects of drinking oxygenated water on proton relaxation times	45
6.2.	Optimisations of T ₁ relaxometry	48
6.3.	Evaluation of relaxometry changes caused by oxygen	51
6.4.	Detection of oxygen <i>in vivo</i>	52
6.5.	Selection of target veins	53
7.	Limitations	55
8.	Conclusion and future work.....	57
8.1.	Conclusion.....	57
8.1.	Future perspective	57
9.	References.....	60
10.	Papers I – III.....	63

Abbreviations and definitions

α	Flip Angle
AR1	1st order Auto Regressive
ATP	Adenosine Triphosphate
B ₁	Excitation pulse in MRI
B ₀	Magnetic density / Magnetic field strength (in Tesla)
C	Concentration
Chi ²	“Goodness of fit”
CoV	Coefficient of Variation
DO	Dissolved Oxygen
ECG	Electro Cardiogram
EPI	Echo Planar Imaging
GI	Gastrointestinal
Hb	Haemoglobin
HODC	Haemoglobine-oxygen dissociation curve
HPV	Hepatic Portal Vein
Htc	Haematocrit in %
IGC	Individual Growth Curve
LL	Look-Locker
LoD	Limit of Detection
mg	Milligram
ml	Millilitre
MLM	Mixed Linear Model
MOLLI	Modified Look-Locker Inversion recovery
MRI	Magnetic Resonance Imaging
MR _r	Magnetic Resonance relaxometry
ms	Milliseconds
M _z	Component of the NMV in the z-direction
NMV	Net Magnetisation Vector
O ₂	Oxygen
pO ₂	Oxygen partial pressure
R ₁	Proton relaxation rate in s ⁻¹

r_1	R_1 relaxivity of a contrast agent in $s^{-1} \cdot mg/L$
R^2	“goodness of fit” for a regression analysis
R_2^*	The transversal relaxation rate in s^{-1}
r_2^*	R_2^* relaxivity of a contrast agent
RF	Radio Frequency
ROI	Region of Interest
SI	Signal intensity
SNR	Signal to Noise Ratio
sO_2	Oxygen saturation
T_1	Proton longitudinal relaxation time in ms
T_{1blood}	T_1 time of blood in ms
T_2^*	Proton T_2^* relaxation time in ms
TE	Echo Time in ms
Ti	Time since inversion in ms
Ti_{inc}	Increments in inversion time in ms
Ti_{max}	Maximum inversion time in ms
Ti_{min}	Minimum inversion time in ms
TR	Repetition time in ms
VCI	Vena Cava Inferior

List of Papers

Paper I.

Precision of T_1 relaxation time measurements in the hepatic portal vein; influence of measurement technique and sequence parameters.

Vatnehol SAS, Hol PK, Bjørnerud A, Amiry-Moghaddam M, Haglerød C, Holck Storås T.

Magnetic Resonance Materials in Physics, Biology and Medicine (2019) 32:359-368

doi.org/10.1007/s10334-018-00731-1

Paper II.

Determination of oxygen r_1 at 3 Tesla using samples with a concentration range of dissolved oxygen.

Vatnehol SAS, Hol PK, Bjørnerud A, Amiry-Moghaddam M, Haglerød C, Holck Storås T.

Magnetic Resonance Materials in Physics, Biology and Medicine (2019)

doi.org/10.1007/s10334-019-00783-x

Paper III.

The effect of drinking oxygenated water assessed by *in vivo* MRI relaxometry

Vatnehol SAS, Hol PK, Bjørnerud A, Amiry-Moghaddam M, Haglerød C, Holck Storås T.

Journal of Magnetic Resonance in Medicine (2020) 52:720-728

doi.org/10.1002/jmri.27104

1. Background

Oxygen uptake in humans is achieved by using specialised organs - the lungs - whose purpose is to facilitate gaseous exchange between the body and the atmospheric air. Under normal situations, the lungs facilitate uptake of adequate amounts of oxygen to the bloodstream. In some situations, however, oxygen uptake from the lungs may not meet the body's demand, and extra oxygen supply may be required.

A potential method of oxygen supply, which has been strongly debated in the literature, is the administration of oxygen through the gastrointestinal (GI) tract in the form of oxygenated liquid. One challenge, when evaluating the feasibility of the uptake of oxygen through the GI tract from a liquid, is how and where to measure the uptake. Although pulse oximetry or blood gas analysis give direct measurements of oxygen content, they are commonly limited to measuring oxygen at a single, peripheral site. Using these techniques to measure oxygen levels in the blood vessels around the GI tract is highly invasive and not suited for human studies.

Because oxygen is a paramagnetic molecule, it has the ability to change two of the basic Magnetic Resonance Imaging (MRI) relaxation parameters, T_1 and T_2^* . Since the effect of oxygen as a relaxation manipulator is small compared to other manipulators, such as contrast agents, it is not certain that the effect of oxygen is clearly visual on conventional T_1 and T_2^* - weighted images. Quantitative measurements of relaxation times, by specific techniques known as MRI-Relaxometry (MR_r) may be able to detect changes in relaxation times at lower oxygen concentrations than is needed for visual detection in conventional images.

In this thesis, I wanted to evaluate if it is possible to detect changes in T_1 or T_2^* relaxation times in the hepatic portal vein (HPV) when drinking water with high oxygen content.

2. Introduction

2.1. Usage and transportation of oxygen

Oxygen is a crucial component for proper metabolic processes in humans. At rest, tissue utilizes 20 – 30 % of the oxygen that is delivered. Factors such as exercise, trauma, and agitation increase the body's oxygen demand, while sedation and hypothermia are known to reduce the demand (1). It should be considered well-known that insufficient oxygen supply may result in drastic consequences, with infarction and death being the most severe.

The main oxygen uptake in humans is through the lungs, from where it is further transported through the body by the bloodstream (2, 3). Oxygen is primarily transported bound to the oxygen carrier haemoglobin (Hb), and only a small fraction of the total oxygen content is freely dissolved in the blood (2). The total amount of oxygen in the blood, and thus oxygen available for the tissue, is, therefore, a product of the concentration of Hb, its saturation (sO_2), and oxygen partial pressure (pO_2). The relationship between sO_2 and pO_2 is given by the haemoglobin oxygen dissociation curve HODC (1, 2), which describes the amount of oxygen available under given physiological conditions. Since both the potential toxicity of high oxygen levels and the potential tissue death caused by low levels are well-known (4) this will not be discussed in detail in this work.

2.1.1. The effect of oxygen on uptake through the gastrointestinal epithelium

Fluids are transported through the epithelial layer of the GI tract to the bloodstream. Under normal circumstances, 9 litres of fluids and 1 kilogram of nutrients are transported over the epithelial layer each day. Transport of such volumes results in high energy demand and thus the need for a steady supply of oxygen. During fasting, however, the blood supply to the

intestines is reduced to such a degree that the intestinal tissue experiences a situation of “physiological hypoxia” (5). The hypoxic situation results in reduced activity in transporters relying on Adenosine Triphosphate (ATP). Although the transport of fluids is a passive non-energy-consuming process, it is dependent on having an ionic gradient over the plasma membrane. To maintain this gradient, ions must be transported over the membrane, primarily using the Na^+/K^+ -ATPase pumps, which requires ATP. Commonly the response to eating is to upregulate the blood flow to the intestine supplying the epithelial cells with increased oxygen, thus allowing an increased transportation rate through the Na^+/K^+ -ATPase pumps (5).

2.1.2. Oxygenated water as an alternative oxygen supply

Under normal conditions the lungs supply adequate amounts of oxygen to the body; however, there are conditions where additional oxygen may be required, such as reduced lung capacity, exposure to high altitudes or under increased metabolic demand. Furthermore, oxygen therapy is commonly used in patients with chronic obstructive lung disease, where the patients breathe medical oxygen gas through a mask. As this therapy may have practical restrictions for the patient, there have been studies on the efficacy of per oral administration of oxygen-enriched water as an alternative oxygen-delivering method. The uptake of oxygen by this method is, however, highly debated in the literature. Although some groups have found markers that may indicate some benefits from oxygenated beverages (6, 7), the majority of research on this topic has been unable to detect any physiological effects (8-13). In a study performed by Forth and Adam from 2001, oxygenated water (Dissolved Oxygen (DO) of 45 – 150 mg/L) was administered orally to anaesthetised rabbits by gastric tubes. Oxygen uptake was measured using direct probe measurements in the stomach, abdominal cavity and HPV. Although 30ml water with DO 80 mg/L gave an increase of 10 mmHg in the HPV, the results were not significant until 30 ml of 150 mg/L DO water was administered, giving an increase

of 14 mmHg. Similarly, in the abdominal cavity, water with DO of 80 and 150 mg/L increased the PO₂ by 17 mmHg and 26 mmHg (14).

More recently Charton et al. (15) used anaesthetised pigs and administered per orally either placebo or oxygenated water produced by two methods. No difference in PaO₂, arteriovenous O₂ difference, arterial pressure or heart rate between the groups was reported. They did, however, find that one of the oxygenated beverages, oxygenation by electrolysis, showed a more stable transcutaneous oxygen pressure.

The two studies above used direct measurements by probes and blood samples to evaluate the effect in animals of drinking oxygenated water. Blood samples have also been used in human studies, such as to test for genotoxicity (16), where no genotoxic effects were found. In 2005, Gruber et al. (17) collected blood samples and concluded that long-term consumption of oxygenated water was safe. In this study, volunteers consumed 500 ml of water with DO of 190.6 mg/L three times daily for 21 days. They found no difference between placebo and oxygen groups in most blood parameters analysed; however, they found a significant increase in ascorbyl radicals and a decrease in natural killer cells in the oxygen group which was not found in the placebo group (17).

There have also been studies that have attempted to find effects in physiological performance after drinking oxygenated water. In these studies, they combined exercise tests and blood samples to evaluate possible effects. In one study from 2006, Leibetseder et al. (8) found differences in lactate and volume of expired O₂ between the control and oxygen groups. They did, however, consider this to be a result of statistical effects rather than a real effect. These findings are in concordance with a recent study by Fleming et al. (6), who examined trained male runners and found a higher lactate clearance rate when drinking oxygenated water during exercise compared to a placebo.

One of the objections against oxygenated water as an alternative oxygen supply is that the amount of oxygen available in oxygenated beverages is low compared to what humans under normal circumstances have available from atmospheric air (12, 13). Furthermore, it is uncertain whether the oxygen will be taken up by the bloodstream, and how this will affect the oxygen levels of the blood. As stated by Willmert et al. (12) it is:

“Highly likely that the cells in the gut consume the extra oxygen. Even if the oxygen did get absorbed, it would be absorbed into venous blood, where it is on it’s (sic) route back to the lungs for oxygenation, not to active muscle. As soon as the blood passes through the lungs for the first time, the blood would either release the extra oxygen at the alveolar membrane, or more likely, not need to pick up as much oxygen in the lung in the process of attaining a typical partial pressure of oxygen”(12).

The alveolar “reset” as described by Willmert et al. (12) is important when considering where to perform a study on oxygen uptake from beverages. Since most previous studies have utilised peripheral pulse oximetry and blood samples (6), the alveolar reset may explain the inability to show a result in these studies. It may, therefore, be more sensible to measure uptake from a more centralised area, as done by Forth & Adams (14). Direct measurements by probes from the portal vein are, however, a highly invasive method and thus not an adequate method for human studies. Non-invasive methods for measuring uptake of oxygen are therefore desirable. Among the methods previously employed in humans, animals and phantom studies is MRI, which is sensitive to the changes in relaxation parameters caused by oxygen (18).

The use of MRI to detect uptake of oxygen from beverages is not a novel idea. Nestle et al. used this approach in 2004 using conventional T₁-weighted images (19). They observed a difference in signal intensity (SI) in oxygenated versus control water in the oral cavity and stomach, and concluded that drinking oxygenated water does lead to an increase in the

available abdominal oxygen. The authors were, however, unable to determine if there was an uptake of oxygen into the bloodstream. There have also been studies on the effects that inhaling oxygen has on the appearance of MRI images. Such imaging experiments are commonly titled MRI oxygen challenges.

The term oxygen challenge has been used by several authors to describe MRI experiments where imaging is performed before, after, and during a change in the available oxygen for the participants. A common setup of this procedure is to equip the patient, or volunteer, with an air tight mask supplied with air where the content of different gases can be controlled. This approach has been used to evaluate oxygen kinetics in several organs and tissue types (20-23). Following the change from room air to 100 % oxygen, there is a modest increase in the venous partial pressure of oxygen from 39 mmHg to 48 mmHg; the arterial partial pressure, however, has a high increase from 98 mmHg to 600 mmHg (20). Since both the O₂ molecule and deoxyhaemoglobin are paramagnetic, their presence will affect the T₁ and T₂* relaxation rates, which can be detected either by visual inspection of conventional images or quantified by using MR_r.

Based on the knowledge that dissolved oxygen is paramagnetic, and that the presence of a paramagnetic substance influences both the transversal and longitudinal relaxation times of tissue water, I hypothesised that an optimised MRI technique could be able to measure minimal differences in relaxation rates as a function of increased DO levels. Since relaxometry, unlike SI, is a quantitative measurement, it is plausible that relaxometry would be able to detect smaller increases in oxygenation levels.

Using T₁-measurements for the detection and quantification of increased oxygen makes sense if the oxygen is present in the bloodstream as DO; however, most of the oxygen absorbed will most likely be immediately bound to the haemoglobin when entering the bloodstream. Whilst deoxyhaemoglobin is paramagnetic, oxyhaemoglobin is diamagnetic. Since paramagnetism

induces strong T_2^* relaxation enhancement, it is plausible that if the oxygen is immediately bound to the haemoglobin, T_2^* mapping may be better suited to detecting an increase in sO_2 than T_1 mapping.

2.2.Aims

The main aim of this thesis was to measure the T_1 and T_2^* relaxation times in HPV blood to compare the relaxometry effects of drinking high DO water with those of drinking control water.

The specific objectives of this thesis were:

- Paper I: To examine the longitudinal stability of different relaxation measurement techniques *in vivo*.
- Paper II: To examine the sequence sensitivity to DO levels for a selection of relaxation measurement techniques, and to estimate the relaxivity (r_1) of oxygen in water.
- Paper III: To perform a double blinded randomised control trial to examine whether per oral administration of oxygenated water affected the relaxation times measured in the HPV blood.

3. Methods

3.1. Methodological considerations

The end goal of this thesis was to evaluate whether MR_r is able to detect changes in relaxation times in HPV following per oral intake of highly oxygenated water. Such an evaluation is dependent on optimised sequences in terms of stability and sensitivity. Based on the planned design of the final study, Paper III, it was decided that the two primary qualities of the sequences were longitudinal stability and sensitivity. There was also a question as to whether the oxygen would be present as dissolved or haemoglobin-bound. Ideally, both T₂* and T₁ mapping sequences should have been optimised, however, due to time constraints, the decision was made to focus on the T₁ sequence.

The first experiment was therefore designed to compare the longitudinal stability of a selection of T₁ mapping sequences. The sequences with the highest stability were subsequently evaluated for their sensitivity in Paper II. Combining the results from Papers I and II allowed for the selection of a potentially optimised sequence, which in combination with a T₂* mapping technique was used in Paper III, to evaluate changes in relaxation times after intake of oxygenated and control water.

The methods for oxygen measurements *ex vivo* and MRI techniques used in this thesis are described below.

3.2.Measurement of oxygen concentrations in water by Winkler titration

Oxygen measurements were performed by using Winkler titration, which is considered to be the most reliable method for determining DO concentrations in water (24, 25).

In brief, the test is performed by adding an excess amount of manganese sulphate and alkaline iodide azide, to create a brownish sediment. The sediment is then liquefied with highly concentrated sulphuric acid, before titration with sodium thiosulphate until the solution turns clear. Starch indicator solution is used to facilitate the titration of sodium thiosulphate. The amount of oxygen can be calculated from the amount of sodium thiosulphate used (26).

Figure 1. Water samples with DO ranging from 0.3 to 21.3 mg/L after the addition of manganese sulphate and alkaline iodide azide. The hue of the samples indicates the amount of oxygen in the sample.

3.3.MRI

3.3.1. Basic Magnetic Resonance Imaging

At its core, MRI consists of using a strong magnetic field (B_0) to align the protons (also known as spins) in a sample to generate a Net Magnetisation Vector (**NMV**) along the z-axis of the magnet. Subsequently, the **NMV** is manipulated using the combination of magnetic gradients, and RF pulses (B_1) to “flip” the **NMV** from the z-axis to the xy-plane. By measuring the **NMV** component in the xy-plane following different combinations of gradients and RF pulses, images highlighting different contrast parameters can be produced. The exact processes concerning how MRI signals are generated and spatially encoded falls outside the scope of this thesis – interested readers are instead referred to some of the many books on the topic which describe these processes in greater detail (27-30).

3.3.2. Spin relaxations, spin-lattice and spin-spin interactions

3.3.2.1. T_1 relaxation, spin-lattice relaxation

Once the **NMV** is located in the xy-plane, spin-lattice interactions cause the **NMV** to return to its stable equilibrium state at a rate of R_1 . The inverse of R_1 , the T_1 relaxation time, is the time from spin excitation, B_1 , to where the **NMV** has recovered 63 % of its original magnetisation along the z-axis (31). The relaxation process is a complex function of the local molecular environment as well as the presence of any paramagnetic species.

3.3.2.2. T_1 of blood

The knowledge of T_1 of blood ($T_{1\text{blood}}$) is necessary for several advanced clinical applications, including arterial spin labelling procedures for non-invasive measurements of tissue perfusion (32). There is, however, great variation of reported $T_{1\text{blood}}$ values and some of the work has been performed on *ex vivo* human blood or in animal models (33-39). The large range of T_1

values reported (32, 38, 40, 41) may be due to variations both in several factors intrinsic to the blood being measured (e.g. haematocrit) and other physiological variables (e.g. sO₂ pO₂, temperature, heart rate and age) as well as factors intrinsic to the measurement techniques being used (e.g. sequence type and field strength).

3.3.2.3. T_2^* relaxation

Simultaneously with T_1 relaxation, T_2^* relaxation will occur. Immediately after the **NMV** is oriented with a component in the xy-plane, all the protons are in phase, thus giving the strongest possible **NMV**. Due to local field inhomogeneity, the spins experience magnetic fields that are slightly different from B_0 , resulting in a difference in Larmor frequencies depending on the spatial position of the spins. This difference in frequencies results in a dephasing and reduction of **NMV** (27, 28, 31). The relaxation rate R_2^* , and its inverse the T_2^* time, are dependent on the magnitude of the inhomogeneity.

3.3.2.4. T_2^* of blood

As with T_1 , there are several factors that contribute to the T_2^* value of blood. For this thesis one of the more interesting is the difference in susceptibility between oxyhaemoglobin and deoxyhaemoglobin where the former is diamagnetic and the latter is paramagnetic. In fully oxygenated blood, the high concentration of oxyhaemoglobin and low concentration of deoxyhaemoglobin results in a match between the susceptibility of intracellular and extracellular space. This match in susceptibility results in increased T_2^* times when compared to low oxygenated blood dominated by deoxyhaemoglobin where the difference in susceptibility is increased (42).

3.3.2.5. Evolution of the NMV

If the relaxation rates described above, and the proton density of a sample are known, then the signal intensity of a given voxel can be estimated using the equations:

$$SI(TR, TE) \propto M_0 (1 - e^{-\frac{TR}{T_1}}) (e^{-\frac{TE}{T_2}}) \quad [1]$$

for spin echo sequence, and

$$SI(TR, TE, \alpha) \propto M_0 \frac{\sin(\alpha) * (1 - e^{-\frac{TR}{T_1}})}{1 - e^{-\frac{TR}{T_1}} \cos(\alpha)} e^{-\frac{TE}{T_2^*}} \quad [2]$$

for spoiled gradient echo sequence.

In these equations, TR, TE and α are the repetition time, echo time and flip angle, respectively. Although T_1 and T_2^* relaxation times are considered to be intrinsic contrast parameters, they may be influenced by the presence of other magnetic, or paramagnetic, substances. This effect is commonly used in MRI in the form of contrast agents. The strength by which a contrast agent affects the adjacent volume is the relaxivity of the agent, denoted r_1 and r_2 .

In terms of T_1 , the effect of the contrast agent can be written as

$$R_1 = R_{1b} + r_1 * C \quad [3]$$

where R_{1b} and R_1 are the relaxation rates with and without the contrast agent, r_1 is the relaxivity, and C is the concentration.

It could be stated that whole blood consists of pure water with the different blood components working as contrast agents to produce the T_1 that is measured in blood. In this model, the contribution of each blood component must be viewed separately.

3.3.3. Quantitative MRI

As stated above one of the strengths of MRI is the ability to produce images with different weightings, where one of the intrinsic relaxation parameters is emphasised. Such images can be considered qualitative in nature, where each pixel is presented with a relative SI compared to the other pixels. It is, however, possible to also acquire quantitative information where the SI of a pixel gives a quantitative, and comparable, value (43). Quantitative MRI techniques, such as diffusion, perfusion, flow and relaxometry are in widespread use for several organs (43-46).

The quantitative method of MRI relaxometry analysis (MR_t), also known as relaxometry — mapping and —measurements, refers to techniques that attempt to quantify and describe the relaxation processes described above. MR_t measurements are based on acquiring several images where one sequence parameter is varied. Assuming that the parameter changed influences the T_1 or T_2^* weighting and that the evolution of the signal is well-behaved, relaxation rates and times can be calculated. Although, in theory, it is possible to measure T_1 with only two images, it is generally accepted that at least five measurement points are necessary (27, 47).

Figure 2 Evolution of SI. Theoretical SI (black circles) and measurements acquired from a volunteer in study 3 (open circles) using the MOLLI 10(5)5 sequence. As evident from the figure, the SI measurements show a good fit but do not align perfectly. The error bars indicate the standard deviation of the SI measurement whilst the dotted line represents the fitted curve of the theoretical measurements.

The description above includes some assumptions that are not necessarily true. Since estimations of T_1 are based on fitting the SI of an ROI or a pixel to the relaxation curve, errors in SI measurement can result in biased T_1 measurements. Such errors may be caused by noise, artefacts, saturation, motion or the readout process itself.

A basic method for T_1 relaxation measurement is inversion recovery sequences where **NMV** is inverted with a 180° pulse at the beginning of the sequence. After a delay time, T_i , a 90° excitation pulse is applied to move the **NMV** into the xy-plane. By repeating this sequence with varying inversion saturation times (T_i), a relaxation curve can be estimated (27). This method requires full relaxation of the magnetisation curve between repetitions, generally accepted to be in the order of $5 \times T_1$. This method is, therefore, time-consuming and not clinically feasible, especially when imaging areas with respiratory or cardiac motion. Alternative methods have therefore been suggested and implemented across vendors.

3.3.3.1. The Look-Locker method

An important milestone in T_1 mapping was the development of the Look-Locker (LL) method, which was suggested by Look and Locker in 1970 (48). Following a 180° inversion pulse, a train of RF pulses is used to initiate a readout of the **NMV**. A signal is measured for each RF pulse, and the signals can be fitted to a relaxation curve. In imaging, this is commonly combined with echo-planar imaging (EPI) readouts allowing faster image acquisition. EPI distortions may be reduced by applying segmented EPI, or by a combination of segmented EPI and transient field echo (TFE-EPI). In summary, the LL method allows the sampling of full or partial k-space for several images with varying T_i for each TR. Figure 3 shows a simplified sequence diagram of a LL-sequence

Figure 3. Simplified sequence diagram of the Look-Locker sequence from Paper I (49). The real relaxation curve is shown as a dotted line, whilst for the observed readout curve, the effect of the readout pulse, is shown as a solid line. Increasing the flip angle α increases the effect on the observed relaxation curve causing an error in T_1 estimates. This effect appears to be mitigated when analysing voxels with through plane flow.

Important parameters for the Look-Locker sequence include the time from the inversion pulse to the first and last readout segments ($T_{\min}$ and $T_{\max}$, respectively), the time between the

readout segments ($T_{i_{inc}}$), and the number of readout segments. The time between two successive inversion pulses is also important, as it should be long enough for sufficient relaxation of the **NMV** to avoid saturation.

One limitation of this sequence is that the magnetisation recovery trajectory is influenced by the readout process, causing a saturation or steady-state effect. Unless accounted for, this could result in errors in T_1 -estimation. To mitigate the saturation effect, low flip angles and high EPI accelerations are used. The low flip angles do, however, result in a reduced SNR. Since blood vessels contain flowing protons that have not experienced a series of RF excitations (only the initial non-selective inversion pulse) higher flip angles can be used in the estimation of T_1 of flowing blood (37, 50).

One further aspect of the LL method is that the images are acquired without regard to cardiac or breathing motion. The technique therefore functions well in areas without physiological motion such as the brain or extremities, but it might not be suitable for evaluating T_1 of areas with organ motion (47).

3.3.3.2. *Modified Look-Locker Inversion recovery (MOLLI)*

MOLLI was proposed by Messroghli et al. in 2004 with the purpose of developing a technique more adapt to measuring T_1 in the moving heart (51). This is achieved by fixing the readout to a specific cardiac phase, and using a flow insensitive single shot balanced technique for image acquisition. This allows for a pixel-wise evaluation of the cardiac muscle.

Figure 4 shows a simplified sequence diagram of a MOLLI sequence.

Figure 4. Simplified sequence diagram of the MOLLI sequence from Paper 1 (49). The sequence is initiated by a 180-degree excitation pulse followed by readout segments separated by one heartbeat. The increment between readout segments ($T_{I_{inc}}$) is then defined as $1/\text{heartrate}$. Readout duration is increased by increasing the number of readouts following one inversion pulse, whilst higher density can be achieved by increasing the number of MOLLI cycles. Complete relaxation between the MOLLI cycles is achieved by adding a recovery phase between the last readout segment of a cycle and the following inversion pulse.

The readout duration is now a result of the number of heartbeats following the inversion pulse. Higher density in readout points can be achieved by adding more cycles to the readout while shifting the timing of the inversion pulse with respect to the cardiac phase. This addition of several cycles results in a MOLLI scheme that describes the number of readouts for each cycle, as well as the recovery phase between them. Care must be taken to allow for sufficient relaxation between the cycles to avoid biased estimations of T_1 . (47)

3.3.3.3. *Effect of technical parameters on T_1 -mapping*

There are numerous technical parameters that may influence the accuracy and precision of T_1 -mapping. For this thesis, only the factors considered in Paper I will be addressed.

3.3.3.3.1. *Readout duration*

The readout duration, $T_{i_{\max}}$, in LL and MOLLI describes the time from the initial inversion pulse to the final readout point. In this thesis, this corresponds to the time between two successive inversion pulses in LL and the duration of the longest cycle in MOLLI sequences.

A general assumption is that the $T_{i_{\max}}$ needs to be 5 times the longest expected T_1 that we want to measure (27, 47). Insufficient $T_{i_{\max}}$ can potentially result in two challenges. Firstly, in a model where only one inversion is used, an insufficient readout time results in less information about the relaxation curve. This might affect the T_1 relaxation times estimated.

Figure 5 visualises this using a MOLLI 10(5)5 as an example. Using the SI from all readout segments, only the first cycle and only the ones included in a 5(3)3 readout result in three different appearances of the curves, as well as three slightly different T_1 estimates.

Figure 5. The effect of readout duration and timing of readout segments. The three variations resulted in three different T_1 estimates although the input SI is equal for all variations. T_1 estimates were: 1844 (original) 1898 (first cycle only) and 1838 (reduced scheme). Note that insufficient time for relaxation might result in the signal being saturated in repeated MOLLI cycles.

Secondly, a more complex situation occurs when multiple repetitions are required, as with multi-shot imaging in LL, or with the addition of more MOLLI cycles. If **NMV** is not allowed to fully relax before the addition of the next inversion pulse, then the resulting measured signal will be compromised by the reduction of **NMV** which affects the calculated T_1 . It should be noted that the rest period in MOLLI allows for a period of undisturbed relaxation, which lessens the saturation effect.

3.3.3.3.2. *Flip angle*

As described in the section regarding LL the readout process affects the relaxation curve, resulting in a saturation effect of the **NMV**. To minimise the saturation effect a train of low flip angles is typically used in the readout process. The use of a low flip angle does result in a decreased SNR by the equation

$$SNR \propto M_z * \sin(\alpha) \quad [4]$$

where M_z is the magnetisation along the z-axis and α is the flip angle.

Whilst the low flip angle is necessary for stationary tissue, high flowing blood does not experience multiple readout pulses. In LL imaging, this effect is evident with less manipulation of the relaxation curve, indicating that a higher flip angle can be used if only the T_1 of flowing blood is of interest (37, 50).

3.3.3.3.3. *Timing method*

Two methods of timing have been used in this work: real-time timing, which utilises the patient or volunteer's electrocardiograph (ECG) signal for triggering, and fixed timing, where a simulated ECG signal is used as a triggering method.

For LL, the choice of timing method affects the $T_{i_{max}}$ - setting a $T_{i_{max}}$ to 5 beats indicates a $T_{i_{max}}$ and TR of 5 beats. In practice, this results in $T_{i_{max}}$ being proportional to the heart rate.

Naturally, this also affects the number of readout time points that is achievable. Although the factors affecting sampling duration and minimum echo spacing are considered to be beyond the scope of this thesis, if all other factors are equal, a reduction in $T_{i_{max}}$ also reduces the number of readout points, since these are given by $T_{i_{max}}$ and $T_{i_{inc}}$.

The same link between $T_{i_{\max}}$ and heart rate exists with MOLLI as with LL. With MOLLI, the link between $T_{i_{\text{inc}}}$ and heart rate is, however, stronger with $T_{i_{\text{inc}}}$ being inversely proportional to heart rate.

Fixed timing, therefore, allows for a standardisation of $T_{i_{\max}}$, $T_{i_{\text{inc}}}$ and number of T_i 's.

3.3.3.4. T_2^* measurements.

For the final Paper in this thesis (Paper III) a multi Fast Field Echo (mFFE) sequence was used to generate images used for T_2^* relaxation time measurements. The sequence is based on a spoiled gradient echo sequence with a TR of 71 ms and a flip angle of 25° . In this sequence 18 echoes with a minimum TE of 1.2 ms and an echo spacing of 4 ms.

3.3.4. Ethical approval

The work was conducted in accordance with the Helsinki Declaration. The studies were approved by the Regional Committees for Medical and Health Research Ethics based on approval 2016/1746/REK sør-øst C, Norway.

Written and verbal descriptions of the studies were given to all volunteers, who signed written informed consent forms before inclusion in the studies. Volunteers were informed about their right to withdraw from the study at any time without providing any reason thereof.

3.3.5. MRI scanner and volunteers

All MRI examinations were performed at the Intervention Centre at Oslo University Hospital, Oslo, Norway using a 3 Tesla Ingenia Scanner from Philips Medical Systems (Best, The Netherlands) equipped with an anterior coil and embedded posterior coil. The same coil setup was used for both volunteer and phantom studies. For acquisitions that were not cardiac triggered, a simulated ECG signal of 60 beats per minute was used.

Volunteers were recruited by placing posters at local hospitals and universities, and by sending e-mails to participants who had previously attended research projects and had

communicated an interest in participating in further studies. For Paper I and III respectively, 10 and 30 healthy male volunteers were recruited. Inclusion criteria were male, age 20 - 35, no regular medications, no known liver or blood disease, and no contraindications against MRI.

3.3.6. MRI scan parameters

Selected parameters of the most central MRI sequences used in the study can be found in Table 1 bellow.

Table 1

Paper	Sequence ^{a)}	TR	TE	Aqu. resolution (mm)	Re. resolution (mm)	Pre Pulses	Scan / Breath hold duration	Inversion times (ms)
1,2,3	MOLLI 10(5)5	Shortest (2.5 ms)	Shortest (1.09 ms)	1.5x1.5x10	0.49x0.49x10	Invert	20 ms / $\approx$ 21 ms	181 ms (1 st shot) 470 ms (2 nd shot) Δ 1000 ms
1,2	MOLLI 5(3)3	Shortest (2.7 ms)	Shortest (1.22 ms)	1.5x1.5x10	0.49x0.49x6	Invert	11 ms / $\approx$ 12 ms	199 ms (1 st shot) 350 ms (2 nd shot) Δ 1000 ms
1,2	Look Locker ^{b)}	Shortest (13 ms)	Shortest (5.8 ms)	1.5x1.5x5	0.49x0.49x5	Invert	16.8 ms / $\approx$ 30 ms	78 ms Δ 199 ms (25 phases)
2	Saturation Recovery	Shortest (13 ms)	Shortest (5.8ms)	1.5x1.5x5	0.49x0.49x5	Saturate	1 min 30 s	38 ms Δ 199 ms (25 phases)
3	mFFE	70 ms	1.17 ms Δ 4 ms (18 echoes)	1.5x1.5x10	0.94x0.94x10		16.8 ms / $\approx$ 17 ms	NA
Table 1 Selected sequences paramaters form the sequences used in this study. ^{a)} Sequences selection based on relevance. ^{b)} Look Locker variation with the lowest CoV, in Paper II a variation with 7° flip angle was also included. Other parameters kept equal.								

In addition to the sequences in Table 1, a phase contrast quantitative flow measurement was performed in Paper 1. Velocity encoding in this sequence was set to 100 cm/s, and 25 phases were collected. TR/TE was set to shortest which gave actual values of $\approx 4.6/3$ ms. The sequence was ECG triggered and acquired during a single breath hold of ≈ 19.4 ms.

In the second study, presented in Paper II, a stationary phantom was constructed. Since this phantom did not include flow, the saturation effect of using a 45° flip angle in a LL sequence would likely be strong. Therefore, in addition to the benchmark sequence and the two sequences with the lowest variation from Paper I a LL with reduced flip angle (7°) was also included. Also, since breath hold was not an issue in this study a saturation recovery sequence was also included in the gadolinium group.

3.3.7. MRI materials

A summary of the volunteers and specimens used in the thesis are presented in Table 2 below:

Study #	Volunteer #	Age Mean (range) in years	Time between sessions Mean (Range) in days
1	10	28 (23 – 32)	23.1 (7 – 73 ¹)
3	30	27 (20 – 34)	13.8 (5 ² – 42)
	Total # samples	Removed samples from non-gadolinium group	Removed samples from gadolinium group
2	88	4 ³⁾	1 ⁴⁾

Table 2. Summary of selected volunteer and specimen data.

¹⁾ Due to scheduling difficulties one participant had 73 days between examinations, removing this outlier for the time range results in a mean of 13.8 range: 7-21

²⁾ In the study protocol, as well as the published paper, the time between sessions was presented as “minimum one week”. During the writing of this thesis a scheduling error was identified with two participants attending with 5 days interval.

³⁾ Error in Winkler Titration n = 4

⁴⁾ Error in gadolinium addition n=1

There were no significant difference between the ages of volunteers in study 1 and study 3 (unpaired samples t-test p = 0.34)

3.3.8. Image analysis

All image analysis was performed in nordicICE (NordicNeuroLab, Bergen, Norway). In Paper I, T₁ relaxation times were estimated by placing an elliptical region of interest (ROI) in the centre of the hepatic portal vein on the raw data images. The ROIs were positioned to include as much of the vein as possible while avoiding any visual artefacts or vessel walls. The measurements were performed three times to compensate for intra-reader variability. T₁ from these three measurements was averaged, and the resulting average was noted as the T₁ of that acquisition.

In Paper II, R_1 and goodness of fit χ^2 maps were calculated, and ROIs were positioned based on the χ^2 maps including as much as possible of the sample while avoiding areas of high values (poor goodness-of-fit) in the χ^2 maps. The ROI position was then copied to the corresponding R_1 maps. As in Paper I, all measurements were performed three times.

In Paper III, T_1 , T_2^* , R_1 and R_2^* maps were calculated. Although the data from the R_1 and R_2^* maps were not included in the study, these maps were used to facilitate ROI placement. ROIs were placed in image planes perpendicular to HPV and vena cava inferior (VCI). All measurements were performed three times. As in Paper I, the ROIs were positioned to include as much of the vessels while avoiding vessel walls and visual artefacts.

While not addressed directly in Paper I, a comparison of placing ROI's in the source, or raw, dataset versus placing the ROI in reconstructed parametric maps was also performed.

3.3.9. Calculations of limits of detection (LOD)

In Paper II limits of detection (LOD) was calculated by using the method described by Shrivastava and Gupta (52) using the equation:

$$LD = \frac{F \times SD}{b}$$

Where b is the slope of the regression line, equal to r_1 in this thesis, SD is the standard deviation of the *in vivo* measurements in study 1 and F is a factor set to 3.3 for LOD calculations. The choice of $F=3.3$ is a stringent approach which results in a close to 100% confidence interval between two measurements. The use of a $F=2.2$ results in a 95% confidence interval which may be more sensible. Using a $F=2.2$ resulted in LOD of 12.3 and 8.5 for the MOLLI 5(3)3 and 10(5)5 respectively. By combining the estimated r_1 values from MOLLI sequences, with the estimated *in vivo* $1/T_1$ values and standard deviation (SD)

collected from study 1 it is possible to calculate the needed change in DO to produce a detectable change in R_1 .

3.3.10. Statistical methods

3.3.10.1. *Paired samples t-test*

In Paper I, the comparison between the measured T_1 and coefficient of variation (CoV) among the test sequences and the reference sequence was performed by using a paired samples t-test.

Paired sample t-testing allows for a comparison of measurement metrics when the same participant is measured several times. Typically, this is performed when measuring a patient's values before and after treatment to evaluate whether there is a significant difference between time points. In Paper I, the reference sequence can be viewed as time point = 0, the change of protocol as treatment, and the test sequence values as time point = 1. Since it is not unnatural to assume that the measured T_1 varied within the same participant between examination days, each pair of examination days was treated as a separate pair, resulting in up to 19 degrees of freedom for the sequence that was included in all examinations.

3.3.10.2. *Linear regression analysis*

To calculate the r_1 of oxygen in Paper II, a simple linear regression analysis was performed. Medical researchers commonly use linear regression to evaluate the ability of one measurement metric (independent) to predict changes of a second metric (dependent) (53). If the two measurement metrics are plotted, then the results from the linear regression can be viewed as a line in the plot described by the equation:

$$(53) \quad y_i = a + bX_i + e_i \quad [5]$$

Here y_i is the value of the dependent for the X_i value of the independent. The intercept and slope of the curve are a and b respectively, and e_i is the error of the measurements.

In an ideal situation, there is a perfect correlation between the dependent and independent metrics. Such a situation would result in the measurements falling perfectly on the regression line and having a correlation coefficient (R^2) of 1. In practice, there is rarely a perfect correlation, and the measurement points are a certain distance from the regression line. The purpose of regression analysis is therefore to find the a and b values that minimise the differences measured by the independent metric and y_i from equ. [5]. This is commonly done by utilising a least-square method (53).

Figure 6. Example of a linear regression plot based on data from Paper II. In this example, the intercept (a) is 0.55, the slope (b) is $4.8 \cdot 10^{-3}$ and regression coefficient (R^2) of 0.911.

3.3.10.3. *Bland Altman plots and analysis*

Bias is a common concern when comparing two quantitative measurement methods. Whilst regression analysis allows for an evaluation of agreement, it is not a good method for comparing comparability. For this reason, Altman and Bland proposed a method in 1983 which has later been termed “Bland Altman plots” (54-56). This method also allows for an estimation of limits of agreement.

In Bland Altman plots, measurement methods are compared by calculating the difference between and mean of two measurement points.

$$Diff_n = Measurement\ 1_n - Measurement\ 2_n \quad [6]$$

$$Mean_n = \frac{Measurement\ 1_n - Measurement\ 2_n}{2} \quad [7]$$

where n is the n'th measurement point. By calculating the mean and standard deviation of $Diff_n$, the bias and limits of agreement can be assessed (bias = mean, limits of agreement = mean $\pm$ 1.96*standard deviation).

This is normally plotted as in the example below.

Figure 7. Example of a Bland Altman plot adapted from (56) with data values from study 2. The red line indicates the mean, and the blue lines indicate limits of agreement. The yellow arrow shows the bias, which is the difference between mean and 0 (green line)

What is considered as an acceptable bias depends on the situation, and is, according to Bland and Altman (55), a matter of judgement. For the data presented in Figure 7, the bias was 0.0066, which could be considered negligible as it is close to 1 % of mean R_1 from the data.

Bland Altman plots can further be used to perform a regression analysis of Diff_n as a function of Mean_n to check for correlation in the bias. In the example above, there was a correlation in the difference, which resulted in a higher difference for lower R_1 values.

3.3.10.4. Mixed Linear Model (MLM)

General linear models assume independence of samples that are not commonly achieved in longitudinal studies. For this reason, other methods, such as individual growth curve (IGC) models, are commonly used to evaluate treatment effects (57). IGC also has several benefits that are relevant to study 3 of this thesis, among them being the benefit that it allows for the fact that number and timing of measurements may vary between individuals, a factor that is important in Paper III. IGC also allows for the comparison of several factors and is considered a more robust analytical model compared to other methods for longitudinal analysis.

An IGC method, the Mixed Linear Model, was set up as defining individuals as a combination of volunteer and water type, the repeated measure as minutes from the end of drinking, and covariance structure as first-order autoregressive (AR1).

4. Results

Visual assessment of the relaxation time evolution in study 3 showed an immediate reduction of T_1 in the control vessel, VCI, a situation that normalized after 14 minutes. Such an effect was not evident in the T_2^* measurements. When evaluating the VCI curves, there appeared to be no visual difference between the oxygen and control groups. In the HPV visual assessment of the curves indicated a small reduction in the T_1 relaxation times following drinking in the control group, followed by an increase. This initial decrease was not observed in the oxygen group and the increase was consistently higher than in the control group. Also for the T_2^* times oxygen group consistently reported higher values after drinking than the control group.

The results from the visual inspection of the relaxation curves were also evident in the MLM analysis where the act of drinking as such was significant for the normalized T_1 values (HPV 1.575% increase $p < 0.001$ VCI 0.738 decrease $p = 0.03$), however, not for the T_2^* values (HPV 1.244% increase VCI 1.428% increase $p = 0.529$). The effects of oxygen were significant for both metrics in the HPV: (T_2^* : 9.603% increase $p = 0.017$ T_1 : 0.697% increase $p = 0.028$). In the VCI the effect of oxygen was not significant, (T_2^* : 4.224 % increase $p = 0.204$ T_1 : 0.272% decrease $p = 0.576$).

T_1 relaxation time measurements in the gastric ventricle following the intake of water showed a significant difference between the oxygen and control group (oxygen: 2956 ms, control: 4217 ms $p < 0.001$). This corresponded well with the DO levels measured by Winkler titration on sample bottles after all experiments were completed which were 69.34 and 7 mg/L for the oxygen and control group respectively. An attempt was also performed to measure the T_2^* times in the gastric ventricle. However, the artefact levels in the bulk water of the ventricle rendered this analysis impossible.

In the baseline measurements of study 3 the average estimated T_1 relaxation times before normalization were 1814 ms in the HPV and 1771 ms in the VCI, similarly the T_2^* measurements of these veins were 46.53 and 37.98 ms respectively. The average estimated T_1 relaxation times was similar for the baseline measurements in study 3 and the ones estimated by the same sequence in study 1 (1771 ms (+/- 17 ms)). In terms of HPV flow, the velocity measurements in study 1 showed a minimum velocity of 12,37 cm/s (range 8,03- 21,17 cm/s), maximum 18,31 cm/s (Range 13,2 – 26,46 cm/s) and mean velocity 15,43 cm/s (range 9,92 – 24 cm/s). With a slice thickness of 5 mm this flow indicated a minimum T_{i-inc} of a 40ms, or 62 ms if the absolute lowest flow is used, would be necessary for complete replacement of saturated blood between excitations.

In the first study of this thesis, presented in Paper I, the four LL variations with a 3 beat readout and one of the MOLLI variations showed a significantly lower estimated T_1 when compared to the reference MOLLI 5(3)3 sequence. In terms of stability, paired samples t-test comparing the CoV between the test sequences and the reference MOLLI variation showed no significant reduction in CoV. When comparing the effect of specific scan parameters using a t-test we found that increasing readout duration i.e. an increased duration between inversion pulses, flip angle and using simulated ECG were on their own significant factors for increasing stability. Comparison of ROI placement method was presented as a part of the supplementary materials of Paper I. In short, there was a good agreement between the estimated T_1 and CoV between the two methods of analysis. Using a paired samples t-test between all sequences evaluated, no significant differences were found either in estimated T_1 (average 1612 ms vs 1597 ms $p = 0.11$) or CoV (3.4 % vs 3.1 % $p = 0.29$).

While none of the sequence variations showed a significant gain in stability compared to the benchmark sequence, the slight reduction in CoV for the Lock Looker with a 5 beat readout

duration and a 45° flip angle and the MOLLI 10(5)5 variation, both using simulated ECG triggering, prompted the use of these two variations in the future studies.

Using linear regression between the relaxation rate ($R_1 = 1/T_1$) of the MOLLI sequences against DO resulted in estimated r_1 of oxygen ranging from $4.18 - 4.8 * 10^{-3} \text{ s}^{-1} * \text{mg}^{-1} * \text{L}$.

When comparing the r_1 between the two MOLLI sequences, and between the gadolinium and non-gadolinium groups, it was found that the MOLLI 10(5)5 sequence in the gadolinium group reported a significantly higher r_1 than MOLLI 5(3)3 ($4.8 * 10^{-3}$ vs. $4.41 * 10^{-3} \text{ s}^{-1} * \text{mg}^{-1} * \text{L}$ $p < 0.001$ but not in the non-gadolinium group. (Paper II, table 2).

Due to high levels of artefacts from the LL with 45° flip angle these images were not suitable for analysis. Further, the LL with a 7° flip angle had a weak correlation between DO and R_1 with the linear regression showing a R^2 of 0.32 and 0.006 in the non and gadolinium group respectively and subsequent estimated r_1 of $2.29 * 10^{-3} \text{ s}^{-1} * \text{mg}^{-1} * \text{L}$ and $6.23 * 10^{-4} \text{ s}^{-1} * \text{mg}^{-1} * \text{L}$. These results are also shown in figure 8 bellow.

Figure 8 Scatter plots showing the correlation between the estimated R_1 by the LL with a 7 degree flip angle and DO with gadolinium added (black pyramids ▲) and without gadolinium (white triangles ▽) compared to the R^2 values from the MOLLI sequences the LL sequence performed poorly with R^2 values of 0.32 and 0.006 in the non – and gadolinium groups, respectively. The estimated r_1 was also substantially lower with estimates ranging from $6.23 \cdot 10^{-3}$ to $2.29 \cdot 10^{-3}$. For this reason no LoD calculations were performed.

The estimated r_1 from the saturation recovery sequence was $4.45 \cdot 10^{-3} \text{ s}^{-1} \cdot \text{mg}^{-1} \cdot \text{L}$ (95 % confidence interval $4.1 - 4.84 \cdot 10^{-3} \text{ s}^{-1} \cdot \text{mg}^{-1} \cdot \text{L}$). Which was not significantly different from those of the MOLLI sequences (MOLLI 10(5)5 $p = 0.174$ and MOLLI 5(3)3 $p = 0.823$).

Average temperature in the phantom container was 22.9°C (Range: $21.5\text{-}24.5^\circ\text{C}$) in the non-gadolinium and gadolinium groups. Since temperature is a known confounder for relaxometry measurements and thus potentially for estimates of r_1 the estimates were also performed on temperature groups. These results are shown in the supplementary materials for Paper II, Figure 1 and Table 1.

5. Summary of Papers

Paper I. Precision of T_1 -relaxation time measurements in the hepatic portal vein: influence of measurement technique and sequence parameters.

In Paper I, we investigated how selected scan parameters affected the stability of T_1 measurements in the HPV. Prior to the inclusion of the volunteers' phantom and literature, studies were used to select a total of twelve versions of LL and MOLLI sequences, including the vendor-suggested MOLLI variation. When constructing the different variations of LL and MOLLI, the parameters that varied were the readout duration and flip angle for LL, and the scheme for MOLLI. Also, for both LL and MOLLI, some variations included sequences that were ECG-triggered, whilst others used an internal simulated ECG signal for triggering. In addition to the T_1 measurement sequences, phase-contrast flow measurements were performed to evaluate the velocity of the blood flow in the hepatic portal vein.

Stability of the sequences was examined using the standard deviation of the normalised T_1 values to calculate a CoV. To compare the measured T_1 and the stability of the sequences against the reference sequence paired samples t-tests was performed both on the average T_1 by scanning day and the CoV. Paired samples t-tests was also performed on the results based on the three different parameters that varied. The results showed that insufficient readout duration resulted in significantly lower T_1 values and that a LL with 45° flip angle and 5 beat readout gave the lowest variation.

In conclusion, we were able to measure the T_1 in HPV with low variation and without pulsatile artefacts.

Paper II. Determination of oxygen r_1 at 3 Tesla using samples with a concentration range of dissolved oxygen

In Paper II, we studied the sensitivity in DO quantification of the two sequences showed to give the best stability based on Paper I. The vendor-suggested MOLLI variation and an LL with 7° flip angle were also included as references. To test the sensitivity, a phantom was constructed with 10 - 12 water bottles containing a DO gradient from ~ 0 to ~ 22 mg/L. The DO levels were confirmed using Winkler titration. Furthermore, gadolinium was added in the final experiments to mimic the T_1 relaxation time of venous human blood. The measured R_1 in each phantom compartment in each experiment was correlated with DO levels from Winkler titration to calculate the r_1 of oxygen. Comparing the results from this study with the results in Paper I, Limit of Detection (LoD) was calculated by methods described elsewhere (52).

The results in Paper II show that DO can be calculated by using MOLLI 10(5)5 for measuring R_1 and assuming an r_1 of $4.8 \cdot 10^{-3}$ mg/L. Due to artefact level and low R^2 values, the LL sequences were removed from the study and not included in the publication.

In conclusion, we were able to find the r_1 of DO in water and found that the MOLLI 10(5)5 scheme had a lower LoD than the vendor-suggested MOLLI 5(3)3 scheme.

Paper III. The effect of drinking oxygenated water assessed by *in vivo* MRI relaxometry

Following the conclusion of study two, a pilot phase was initiated to compare the results from MOLLI 10(5)5 and LL with 45° flip angle to evaluate the effects that drinking oxygenated water has on relaxation times. After the pilot phase, a MOLLI 10(5)5 non-ECG-triggered and a multi Fast Field Echo (mFFE) sequence was selected for scanning. The HPV was the target vein, with the VCI serving as a control. Images were acquired in planes angulated perpendicular to the HPV and VCI.

After baseline, the volunteer was asked to drink two bottles (approx. 700 ml) of the test water. In the case of motion between baseline and measurement scans, scan planes were re-adjusted. After drinking the volunteers were scanned for 30 minutes.

MLM showed that there was a significant difference in the T_1 and T_2^* times between the oxygen and control water groups in HPV, but not in VCI. In HPV, there was an immediate increase in T_2^* times after drinking followed by a gradual decrease. The increase was higher with the intake of oxygenated compared to control water, and the increased signal also persisted longer. The same trend was seen for the T_1 relaxation times, with the oxygenated water showing a more rapid increase in relaxation times, as well as reaching a higher level than the control group. Interestingly, there appears to be an immediate decrease in T_1 relaxation times for the control group that is not observed for the oxygen group. In VCI, there was no significant effect of drinking either control or oxygenated water, and, although the curve of the oxygen water appeared to be subjectively higher than the control water, this effect was not significant. Although there were no differences in the T_1 relaxation time between control and oxygenated water in the VCI, there was an immediate and strong decrease in relaxation times following drinking, with a return to baseline after 12-14 minutes.

6. Discussion

6.1. Effects of drinking oxygenated water on proton relaxation times

This thesis showed that there is a difference in relaxation time response in the HPV after drinking control water compared to water with increased levels of oxygen.

The increase in T_2^* can be explained by the conversion of the paramagnetic deoxyhaemoglobin to the diamagnetic oxyhaemoglobin - i.e. an increase in sO_2 . It may also be explained as a function of increased DO, since the presence of a paramagnetic species, such as oxygen, may lead to a susceptibility matching (42). This potential susceptibility matching occurs if the increased oxygen levels lead to a reduction in susceptibility differences between the intra- and extra-cellular compartments. Since the T_2^* relaxation decreases with decreasing susceptibility differences, it can be assumed that increased DO will lead to increased T_2^* relaxation times.

However, the increase in T_1 relaxation times in the oxygen group can not be explained by increased DO. T_1 relaxation measurements in the HPV showed a significant difference between the oxygenated water and the control water, where the T_1 times when drinking oxygen increased faster and to a higher level than the control water. Some previous work suggests that it is compatible with a reduced relaxation effect from deoxyhaemoglobin as this is converted to oxyhaemoglobin (58). There was, however, a similar, although smaller increase in T_1 after drinking control water, which would presumably not change oxygenation status of haemoglobin.

Absorption of ingested water into the bloodstream may lead to a dilution effect compatible with an increase in both T_1 and T_2^* relaxation times. This effect could potentially be caused by both water types. The increased values in the oxygen compared to the control group can be

described by the assumption that the excess oxygen increases the absorption rate of water through the intestinal epithelial cells (5). Whilst water transport from the GI tract to the vessels is a passive process, some of the water transport is facilitated on an ionic gradient between the vessels and the epithelial cells. Although we found no literature on the effects of high oxygen levels, it is well known that low oxygen levels in the epithelial cells result in a reduced function of the Na^+/K^+ ATPase pumps, which actively transport the ions in and out of the cells (5). It is therefore plausible that increased levels of oxygen may result in higher production of ATP and thus a stronger transport of the ions. If such a process is indeed the case, then the increased T_1 relaxation times can be attributed to increased dilution of HPV blood.

The higher relaxation times indicate that the presence of oxygen in beverages increased the water absorption from the intestines, increased sO_2 or a combination of both. Previous studies in both humans and animals, using blood samples, physical tests, and direct measurements have shown varying results. In rabbits the administration of oxygenated water gave an increase in pO_2 which correlated with the DO levels of the water administered (14). The findings from Forth and Adams (14) are not in agreement with our results, as an increase in pO_2 should decrease the measured T_1 relaxation time. While our results are compatible with an increase in water absorption as the main factor for changes in relaxometry values, we should not to rule out that some of the oxygen is also transported across the epithelium and into the bloodstream, increasing the sO_2 , but this effect was not large enough to dominate the measured relaxation time changes. In previous studies showing increases in sO_2 peripheral measurements were performed and their results are therefore not directly comparable to our findings. Comparing the findings to the increased lactate clearance found previously (6, 7) is also difficult as no blood sample analysis of this kind was performed in our study.

Whether the observed larger change in T_1 and T_2^* after drinking oxygenated water was due to conversion of deoxy- to oxy-haemoglobin or due to an increased water absorption through the intestinal epithelial cells (5) facilitated by the excess oxygen was not possible to decide from our data. Further studies could answer this. One could e.g. add a dynamic measurement of water in the gastrointestinal tract to see if the rate of absorption were increased by higher water oxygenation.

Whilst there were no differences between water types in the VCI in Paper III there was an immediate decrease in the T_1 relaxation times, before the relaxation times returned to baseline. This effect may be caused by the process known as intravascular haematocrit layering and also by the experimental setup. During scanning, the patient was asked to lay in a supine position on the scan table, and, when a body remains at rest, slow-flowing blood may experience a partial separation of plasma and haematocrit in the peripheral veins (59). Since the plasma has a higher flow than the haematocrit the blood of the VCI is diluted, and T_1 relaxation times are increased. The volunteers were allowed to sit while drinking the experimental water, causing the venous blood to mix and resulting in reduced T_1 relaxation. When the volunteers then lay down, the intravascular haematocrit layering continued, and the T_1 relaxation times increased to the level during baseline.

6.2.Optimisations of T_1 relaxometry

Only two methods for T_1 measurements were evaluated in this thesis: LL and MOLLI. Other techniques do exist, such as variable flip angle and saturation recovery, but these techniques were not included in this thesis due to time limitations and lack of access to optimized sequence protocols on our system.

The sequence found to have the highest stability in Paper I was a LL variation with increased (45°) flip angle, fixed timing, and a readout duration of 5 beats. In Paper I, the difference between using a fixed timing vs ECG-triggered timing, resulted in average times between inversions of 5000 ms, and 4900 ms respectively. Assuming that the measured average T_1 of the reference sequence used is correct (1759 ms) the minimal readout duration should be 8795 ms. From this, it is evident that the optimised LL sequence violates the minimum readout duration assumption. It did, however, give T_1 measurements that were in accordance with the reference sequence, and, as it had the lowest CoV, it was recognised as a potential optimal sequence. A common trait between the sequences that estimated lower T_1 values than the reference sequences is that the time between inversion pulses, or readout time, was shorter than the sequence variations that achieved similar estimated T_1 's as the reference sequence. Whilst none of the sequences evaluated in Paper I showed a significant reduction in CoV compared to the vendor suggested option, it should be noted that CoVs of 2 % and 1.29 % (reference and optimised sequence respectively) indicates a measurement error of 35.18 ms and 22.69 ms. Assuming a total blood volume of 4 L, a DO of 80 mg/L and an intake volume of 0.66 L, an r_1 of oxygen of 4.8 mg/L/s, and a global and immediate uptake then the maximum change in T_1 would be equal to 63.36 ms. With the small potential T_1 - relaxation changes expected following intake of oxygenated water, a small increase in both sensitivity and precision of the measurement technique could thus be of major importance.

Naturally, the above calculations are an oversimplification, as it does not take into account the potential reset function of the lungs, the potential dynamic absorption of oxygen, or the combination of other effects than increased DO from drinking oxygenated water.

One further assumption for this sequence is that the flow velocity in HPV is high enough to allow for the use of increased flip angles. Comparing the spacing of the segments in the LL sequences, with the slice thickness and the minimum velocity, it is evident that the spacing is sufficient to allow for a complete replacement of spins within the slice, thus allowing for the use of high flip angles. The use of high flip angles in the LL methods does result in some limitations in the usage of the sequence. Although the blood flow through the hepatic portal vein is high enough to allow for a 45° flip angle, spins in the stationary tissues do not refresh. Therefore, this flip angle is probably not suitable if the relaxation times of stationary tissues are of interest.

Whilst precision is of great importance when performing longitudinal studies such as those presented in Paper III, the sensitivity of the sequences to minute changes is of equal importance. When selecting sequences for Paper II it was decided to include the same reference sequence as in Paper I, since this is a well-established vendor-suggested sequence and the sequence with the lowest variation. In addition, two sequence variations were included: a LL with a reduced flip angle to minimise saturation in the phantom, and the most stable MOLLI variation, since this sequence fulfilled the requirements of allowing a $5 \times T_1$ relaxation between pulses. Although scanned, the sequence with lowest CoV from Paper I had an artefact level that rendered it unusable when performed on a stationary phantom. A flow-phantom, either with or without pulsatile flow mimicking the blood flow, could have strengthened the study, as it would be more similar to a clinical setting. It was, however, decided that building such a phantom would result in a loss of control of the oxygen levels, and reduce the number of different samples we were able to examine in each experiment. To

mitigate the saturation challenge scans where also performed using a LL with a reduced flip angle (7°). Unfortunately, as presented in the results section, the quality of the regression analysis for this sequence was poor. For this reason the results of the LL sequences were not included in the published paper.

In a review by Taylor et al. in 2016 several techniques for T_1 mapping were evaluated. When comparing MOLLI and LL they both showed good accuracy and precision. LL does, however, have scope for higher potential temporal or T_1 resolution (47). The choice of not using the LL sequence for Paper III could, therefore, appear counter-intuitive, since this sequence gives the lowest variation, and has a potential for higher T_1 resolution. This was, however, discussed after a pilot phase for Paper III, where MOLLI 10(5)5 was able to detect a significant and consistent change in T_1 following administration of oxygenated and control water, whilst LL was unable to show this change. A further reason to use the MOLLI sequence in Paper III is the differences in breath-hold time, which were shorter for the MOLLI sequence, allowing higher compliance among the volunteers. The inability for LL to detect the same changes as MOLLI 10(5)5 remains unexplained. While this issue was not specifically addressed in this thesis, it is possible to hypothesize that the insufficient time between the inversion pulses gives a saturation effect which limits the sensitivity of the sequence. It should be noted, that although not significant, the LL with lowest CoV did indeed give lower estimated T_1 values than the reference sequence. Initially, we attributed the poor linear regression results of the LL sequence to be primarily a result of the challenges derived from a stationary phantom. Taken together with the results from the pilot study however, it may indicate that increasing the readout duration, i.e. the time between inversion pulses, might have mitigated this challenge.

Since the quality of results utilising MRI depends on the sequences used, it is important to recognise that work must be done to optimise the protocols for research and clinical usages. In particular as new sequences are continuously developed and become commercially available.

6.3. Evaluation of relaxometry changes caused by oxygen

In Paper II we estimated the r_1 relaxivity of DO. If r_1 is known, the difference in the DO of two water samples can be calculated by the equation

$$\Delta c = \frac{R_1 - R_{1_0}}{r_1} \quad [8]$$

where Δc is the difference in DO concentration and R_1 and R_{1_0} are measured T_1 relaxation rates in two samples with different DO and r_1 is the relaxivity of oxygen. While the r_1 of oxygen is a physical constant and thus not dependent on scan parameters, as evident from the results presented in Paper II, the choice of scan protocol affects r_1 estimations. Consequently, if R_1 measurements are to be used to estimate concentration changes, application of sequence specific r_1 may improve accuracy.

Although the calculations of oxygen concentrations presented above is a straightforward approach in water phantoms, two aspects make it more difficult in an *in vivo* setting. Firstly, it is not known if all the oxygen will be available as free oxygen, or haemoglobin-bound, or as a combination. It is reasonable to assume that most of the oxygen will be bound to haemoglobin, resulting in an increase in sO_2 and a stable level of free oxygen. Secondly, and different from *in vitro* situations, the uptake of oxygen *in vivo* is likely to be continuous.

The latter aspect of timing is potentially easy to compensate for, as the T_1 signals could, in theory, be analysed as a perfusion sequence with DO as a contrast agent. The former is more difficult to compensate for. Whilst decreased T_1 could be used to calculate the total increase

in DO, and decreased T_2^* can give an indication of increased sO_2 , the negative effects on T_1 from oxyhaemoglobin make the calculations difficult. For this reason, and because of the apparent complex behaviour of the relaxation time seen in the study, it was decided to not attempt a direct quantification of oxygen uptake.

6.4. Detection of oxygen *in vivo*

The concept behind this thesis is that changes either in levels of DO, or haemoglobin-bound oxygen are measurable by changes in T_1 and T_2^* . A core aspect is whether additional oxygen will remain in a dissolved phase or bind itself to haemoglobin. During the initial planning of this study, no information regarding this was found.

As discussed more in depth in section 6.5, increasing either the DO content or the sO_2 should result in a reduction of T_1 or an increase in T_2^* , respectively. Whilst in the most simplistic form, detection of increased O_2 could be achieved by visually inspecting T_1w and T_2^*w images for a change in SI, similar to what is done when identifying lesions of contrast agent uptake, by comparing before and after images a change in uptake may be evident. This method would be limited to visually discriminated changes. A strong limitation of this method is that the change needs to be very significant to produce a visually detectable difference.

A more robust method would be to measure the SI in the area of interest, and to compare this value between before and after images. To compensate for between-scan changes in SI, a reference tissue could also be measured and used for normalisation. Such an approach has been used in a recent study of gadolinium retention in the brain (60) and allows for a semi-quantitative method where more minute changes can be detected. SI measurements can however, only measure the net effect of all underlying relaxation time changes, and cannot separate between the T_1 - and T_2^* - relaxation effects.

In this work MR relaxometry was therefore performed, whereby T_1 - and T_2^* - relaxation effects can be separately quantified, thereby enabling a more nuanced evaluation of oxygenation effects. Quantitative analysis is challenging as a high level of stability and precision is required in order for the method to be useful. A further limitation for our application was that the baseline values between the participants differ because of differences in their normal blood values. The use of an adequately long baseline was therefore necessary. Because the duration of which a volunteer could be asked to attend a scanning session is limited, and because prolonged sessions may result in difficulties with continuous breath-hold imaging, a compromise was made for Paper III to include eight baseline scans.

The 95 % confidence interval LOD calculated in Paper II for the MOLLI 10(5)5 sequence indicates that an 8.5 mg/L increase in oxygen is required to detect a change in R_1 . This calculation assumes that there are no other effects present than an increase in DO and is, as such, based on a simplified model. A more accurate model may be to use blood samples with varying levels of oxygenation and to use a phantom with flow. Due to technical limitations with regards to the control of oxygen levels, this was, however, not attempted.

6.5. Selection of target veins

As it is likely that access oxygen is consumed in the gut, or is returned to normal values after passing through the lungs (12), measurements of oxygen uptake should be performed in areas adjacent to the intestines, or the vessels draining them. Since it has previously been shown in rabbits that increased oxygen levels are measurable in the HPV (14), we considered this vein to be potential target. Also, since it is not impossible that relaxation times may be affected by the presence of large volumes of water in the stomach, a control vein was also needed. Since

the VCI does not receive blood from the intestines until after the level of the portal vein, this vein was considered as a potential control.

Whilst it could be argued that other vessels closer to the intestines may be better targets, the stable flow, and the large lumen of the HPV resulted in this level being selected. Similarly, the size and flow of the VCI were also suitable for this study.

7. Limitations

Prompted by the initial belief that excess oxygen would be transported as DO to the bloodstream, a significant amount of the work in this thesis went into the optimisation of T_1 measurements compared to T_2^* optimisation. Since DO has a known effect on T_1 , it was natural to assume that T_1 relaxation time measurements would be a suitable technique for detecting an increase in total oxygen content in HPV blood. As the study progressed, however, it became clear that the absorbed oxygen in the GI tract may also be bound to the haemoglobin, as dictated by the HODC. The T_2^* measurements were therefore included in Paper III, as this technique is sensitive for the changes in blood oxygenation levels.

A further limitation is that, in all Papers, a single reader was used for ROI placement and analysis. The author of this thesis did all readings. Being a radiographer, the reader was therefore initially trained by two radiologists to ensure correct and confident ROI placements. Unfortunately, the single reader approach results in an inability to compare the inter-reader variability of the sequences. To compensate for the intra-reader variability, all ROI placements were performed three times and an average of these three placements was used. The number of volunteers in both Paper I and Paper III is a limiting factor as it affects the statistical power. In future studies, a larger cohort should ideally be used to strengthen the confidence in the results. This may especially be true for studies such as presented in Paper I, as a good validation of sequences is crucial for the quality of subsequent studies.

In Paper II, there is also the possibility that the addition of gadolinium could introduce uncertainties in the R_1 measurements, which could result in incorrect LoD and r_1 calculations. Also in Paper II, Winkler titration was used as a gold standard for oxygen content - although this technique is well established, the manual processes used in this study could result in some uncertainties in DO measurements. The estimates for r_1 was also performed in a stationary

phantom consisting of pure water, a situation that is not directly comparable to *in vivo* settings. Ideally a phantom consisting of human blood equivalent, where flow velocities and pulsation could be controlled, or an *in vivo* measurement of blood at different states of oxygenation should have been performed. This was, however, found to be too difficult to be achieved in the scope of this doctoral project.

In Paper III, there is the possibility that choosing other vessels as controls or including other measurement sequences may have given a greater opportunity to both detect and quantify the relaxometry effects of drinking oxygenated water. A further limitation of Paper III is that there is an assumption that the initial decrease in T_1 times in the VCI is a result of intravascular haemocrit layering. This hypothesis was not tested by performing the experiment without drinking, and this should be tested in future research.

8. Conclusion and future work

8.1. Conclusion

Although the exact physiological effect of drinking oxygenated water remains unknown, it appears clear that the DO levels of ingested beverages do affect HPV blood composition in a way detectable by MR_r in healthy, fasting, male volunteers, presumably by increasing the absorption rate of water. We cannot, however, rule out that there is a mix of several effects, and that the oxygen content of the HPV is increased in combination with the increased water absorption. Drinking oxygenated water produced a higher and more rapid increase in T_1 and T_2^* relaxation times in the HPV than the increases produced by drinking control water. This difference was not observed in the VCI.

To our knowledge, this is the first study where MR_r is used to evaluate oxygen uptake from per oral administrated water, and the first to suggest an increase in water absorption caused by oxygenated water. This thesis' contribution to science is therefore to provide evidence suggesting that the human body is able to utilize oxygen from beverages, and that this effect is measurable using optimized MRI techniques.

8.1. Future perspective

As shown, there is a need for further studies in order to elucidate the nature of the results observed in this work. In this thesis, evidence is presented that water absorption in the human gut is increased when drinking water with a high oxygen content. One possible method for evaluating this could be to perform a study similar to what is presented in Paper III, but also including a highly T_2 -weighted sequence with sufficient resolution to estimate the water volume in the intestines. The use of such a sequence could potentially confirm, or debunk, our hypothesis that increased DO increases water absorption.

Since healthy males were used as volunteers in these studies, the effects in females and in patients are not addressed. A further aspect is, therefore, to investigate if increased uptake is evident in patients with reduced lung or GI functions, in order to evaluate potential beneficial health aspects.

If the assumption that high levels of oxygen in water increases water uptake is correct, it should also be of interest to examine if DO in combination with other nutrients result in an increased, or more rapid, uptake of these nutrients. Such a study could be performed both in volunteers, and in selected patient groups. Since the effect observed in Paper III is small, future studies could potentially benefit from increasing the number of participants.

Also, in Paper III the effect of haemocrit layering should be evaluated further, a potential study design could be to include a “third arm” in the study where the experiment is performed both with control and oxygenated water, and without drinking, but where the process is simulated.

An additional aspect of this thesis is that it sheds light on the necessity for image optimisation when using MRI as a research tool. In theory, it would have been possible to perform the study in Paper III without image optimisation and using the vendor-suggested sequences, or sequences optimised from other vendors. However, it is our belief that the pilot phases and image optimisations are crucial when planning MRI research studies. A further benefit of the optimisation of the techniques in this thesis for portal vein relaxometry is that these techniques could potentially be used in future clinical and research situations where quantification and detection of substance uptake are of interest. The optimised sequences presented in this thesis should, however, be viewed with scrutiny, as further technical improvements and sequence developments will render them obsolete in time.

9. References

1. Dunn J-O, Mythen M, Grocott M. Physiology of oxygen transport. *BJA Education*. 2016;16(10):341-8.
2. Collins JA, Rudenski A, Gibson J, Howard L, O'Driscoll R. Relating oxygen partial pressure, saturation and content: the haemoglobin-oxygen dissociation curve. *Breathe (Sheff)*. 2015;11(3):194-201.
3. Ganong WF. Review of medical physiology. 19th ed. Stamford, Connecticut: Appleton & Lange; 1999. xiv, 754 p. p.
4. Wagner PD. The biology of oxygen. *Eur Respir J*. 2008;31(4):887-90.
5. Ward JB, Keely SJ, Keely SJ. Oxygen in the regulation of intestinal epithelial transport. *J Physiol*. 2014;592(12):2473-89.
6. Fleming N, Vaughan J, Feeback M. Ingestion of oxygenated water enhances lactate clearance kinetics in trained runners. *Journal of the International Society of Sports Nutrition*. 2017;14:9.
7. Jenkins A, Moreland M, Waddell TB, Fernhall B. EFFECT OF OXYGENIZED WATER ON PERCENT OXYGEN SATURATION AND PERFORMANCE DURING EXERCISE. *Medicine & Science in Sports & Exercise*. 2001;33(5):S167.
8. Leibetseder V, Strauss-Blasche G, Marktl W, Ekmekcioglu C. Does Oxygenated Water Support Aerobic Performance and Lactate Kinetics? *Int J Sports Med*. 2006;27(03):232-5.
9. McNaughton LR, Kenney S, Siegler J, Midgley AW, Lovell RJ, Bentley DJ. The effect of superoxygenated water on blood gases, lactate, and aerobic cycling performance. *International journal of sports physiology and performance*. 2007;2(4):377-85.
10. Piantadosi CA. "Oxygenated" water and athletic performance. *British Journal of Sports Medicine*. 2006;40(9):740-1.
11. Wing-Gaia SL, Subudhi AW, Askew EW. Effects of purified oxygenated water on exercise performance during acute hypoxic exposure. *International journal of sport nutrition and exercise metabolism*. 2005;15(6):680-8.
12. Willmert N, Porcari JP, Foster C, Doberstein S, Brice G. The effects of oxygenated water on exercise physiology during incremental exercise and recovery. *JEPonline*. 2002;5(4):6.
13. Hampson NB, Pollock NW, Piantadosi CA. Oxygenated water and athletic performance. *JAMA*. 2003;290(18):2408-9.
14. Forth W, Adam O. Uptake of oxygen from the intestine-- experiments with rabbits. *European journal of medical research*. 2001;6(11):488-92.
15. Charton A, Péronnet F, Doutreleau S, Lonsdorfer E, Klein A, Jimenez L, et al. Effect of administration of water enriched in O(2) by injection or electrolysis on transcutaneous oxygen pressure in anesthetized pigs. *Drug Design, Development and Therapy*. 2014;8:1161-7.
16. Speit G, Schutz P, Trenz K, Rothfuss A. Oxygenated water does not induce genotoxic effects in the comet assay. *Toxicol Lett*. 2002;133(2-3):203-10.
17. Gruber R, Axmann S, Schoenberg MH. The influence of oxygenated water on the immune status, liver enzymes, and the generation of oxygen radicals: a prospective, randomised, blinded clinical study. *Clin Nutr*. 2005;24(3):407-14.
18. Nestle N, Baumann T, Niessner R. Oxygen determination in oxygen-supersaturated drinking waters by NMR relaxometry. *Water Research*. 2003;37(14):3361-6.
19. Nestle N, Wunderlich A, Nussle-Kugele K. In vivo observation of oxygen-supersaturated water in the human mouth and stomach. *Magn Reson Imaging*. 2004;22(4):551-6.
20. O'Connor JPB, Jackson A, Buonaccorsi GA, Buckley DL, Roberts C, Watson Y, et al. Organ-specific effects of oxygen and carbogen gas inhalation on tissue longitudinal relaxation times. *Magnetic Resonance in Medicine*. 2007;58(3):490-6.

21. Dani KA, Moreton FC, Santosh C, Lopez R, Brennan D, Schwarzbauer C, et al. Oxygen challenge magnetic resonance imaging in healthy human volunteers. *J Cereb Blood Flow Metab.* 2017;37(1):366-76.
22. O'Connor JPB, Robinson SP, Waterton JC. Imaging tumour hypoxia with oxygen-enhanced MRI and BOLD MRI. *The British Journal of Radiology.* 2019;92(1096):20180642.
23. Uematsu H, Takahashi M, Hatabu H, Chin C-I, Wehrli SL, Wehrli FW, et al. Changes in T1 and T2 Observed in Brain Magnetic Resonance Imaging With Delivery of High Concentrations of Oxygen. *Journal of Computer Assisted Tomography.* 2007;31(5):662-5.
24. Shriwastav A, Sudarsan G, Bose P, Tare V. Modification of Winkler's method for determination of dissolved oxygen concentration in small sample volumes. *Analytical Methods.* 2010;2(10):1618-22.
25. Shriwastav A, Sudarsan G, Bose P, Tare V. A modified Winkler's method for determination of dissolved oxygen concentration in water: Dependence of method accuracy on sample volume. *Measurement.* 2017;106:190-5.
26. Moen I, Ugland H, Strömberg N, Sjöström E, Karlson A, Ringstad L, et al. Development of a novel in situ gelling skin dressing: Delivering high levels of dissolved oxygen at pH 5.5. *Health Science Reports.* 2018;1(7):e57.
27. McRobbie DW. MRI from picture to proton. 2nd ed. Cambridge, UK ; New York: Cambridge University Press; 2007. xii, 394 p. p.
28. Hashemi RH, Bradley WG, Lisanti CJ. MRI: The Basics: Lippincott Williams & Wilkins; 2004.
29. Edelman RR. Clinical Magnetic Resonance Imaging: Saunders; 2006.
30. Bjørnerud A. The physics of Magnetic Resonance Imaging Fys-KJM 4740. 2008.
31. Mugler JP, III. Basic Principles. In: Edelman RR, Hesselink JR, Zlatkin MB, Cruess JV, editors. *Clinical Magnetic Resonance Imaging.* 1. Philadelphia: Saunders Elsevier; 2006. p. 23-57.
32. Muir ER, Zhang Y, San Emeterio Nateras O, Peng Q, Duong TQ. Human vitreous: MR imaging of oxygen partial pressure. *Radiology.* 2013;266(3):905-11.
33. Dobre MC, Uğurbil K, Marjanska M. Determination of blood longitudinal relaxation time (T1) at high magnetic field strengths. *Magnetic Resonance Imaging.* 2007;25(5):733-5.
34. Wu W-C, Jain V, Li C, Giannetta M, Hurt H, Wehrli FW, et al. In vivo venous blood T1 measurement using inversion recovery true-FISP in children and adults. *Magnetic Resonance in Medicine.* 2010;64(4):1140-7.
35. Zhang X, Petersen ET, Ghariq E, De Vis JB, Webb AG, Teeuwisse WM, et al. In vivo blood T1 measurements at 1.5 T, 3 T, and 7 T. *Magnetic Resonance in Medicine.* 2013;70(4):1082-6.
36. Kalavagunta C, Michaeli S, Metzger GJ. In vitro Gd-DTPA relaxometry studies in oxygenated venous human blood and aqueous solution at 3 and 7 T. *Contrast Media & Molecular Imaging.* 2014;9(2):169-76.
37. Qin Q, Strouse JJ, van Zijl PC. Fast measurement of blood T1 in the human jugular vein at 3 Tesla. *Magn Reson Med.* 2011;65(5):1297-304.
38. Varela M, Hajnal JV, Petersen ET, Golay X, Merchant N, Larkman DJ. A method for rapid in vivo measurement of blood T1. *NMR in Biomedicine.* 2011;24(1):80-8.
39. Shimada K, Nagasaka T, Shidahara M, Machida Y, Tamura H. In vivo measurement of longitudinal relaxation time of human blood by inversion-recovery fast gradient-echo MR imaging at 3T. *Magn Reson Med Sci.* 2012;11(4):265-71.
40. Liu P, Chalak LF, Krishnamurthy LC, Mir I, Peng SL, Huang H, et al. T1 and T2 values of human neonatal blood at 3 Tesla: Dependence on hematocrit, oxygenation, and temperature. *Magn Reson Med.* 2016;75(4):1730-5.
41. Piechnik SK, Ferreira VM, Lewandowski AJ, Ntusi NA, Banerjee R, Holloway C, et al. Normal variation of magnetic resonance T1 relaxation times in the human population at 1.5 T using ShMOLLI. *Journal of Cardiovascular Magnetic Resonance.* 2013;15(13):11.
42. Bjørnerud A, Briley-Saebo K, Johansson LO, Kellar KE. Effect of NC100150 injection on the (1)H NMR linewidth of human whole blood ex vivo: dependency on blood oxygen tension. *Magn Reson Med.* 2000;44(5):803-7.

43. Pierpaoli C. Quantitative brain MRI. *Top Magn Reson Imaging*. 2010;21(2):63.
44. Calamante F. Perfusion MRI using dynamic-susceptibility contrast MRI: quantification issues in patient studies. *Top Magn Reson Imaging*. 2010;21(2):75-85.
45. Jones DK. Precision and accuracy in diffusion tensor magnetic resonance imaging. *Top Magn Reson Imaging*. 2010;21(2):87-99.
46. Seraphim A, Knott KD, Augusto J, Bhuva AN, Manisty C, Moon JC. Quantitative cardiac MRI. *J Magn Reson Imaging*. 2019.
47. Taylor AJ, Salerno M, Dharmakumar R, Jerosch-Herold M. T1 Mapping: Basic Techniques and Clinical Applications. *JACC Cardiovasc Imaging*. 2016;9(1):67-81.
48. Look DC, Locker DR. Time Saving in Measurement of NMR and EPR Relaxation Times. *Review of Scientific Instruments*. 1970;41(2):250-1.
49. Vatnehol SAS, Hol PK, Bjørnerud A, Amiry-Moghaddam M, Haglerød C, Storås TH. Precision of T1-relaxation time measurements in the hepatic portal vein: influence of measurement technique and sequence parameters. *Magnetic Resonance Materials in Physics, Biology and Medicine*. 2019.
50. Li W, Liu P, Lu H, Strouse JJ, van Zijl PC, Qin Q. Fast measurement of blood T1 in the human carotid artery at 3T: Accuracy, precision, and reproducibility. *Magn Reson Med*. 2016;77(6):7.
51. Messroghli DR, Radjenovic A, Kozerke S, Higgins DM, Sivananthan MU, Ridgway JP. Modified Look-Locker inversion recovery (MOLLI) for high-resolution T1 mapping of the heart. *Magnetic Resonance in Medicine*. 2004;52(1):141-6.
52. Shrivastava A, Gupta V. Methods for the determination of limit of detection and limit of quantitation of the analytical methods. *Chronicles of Young Scientists*. 2011;2(1):21-5.
53. Zou KH, Tuncali K, Silverman SG. Correlation and simple linear regression. *Radiology*. 2003;227(3):617-22.
54. Altman DG, Bland JM. Measurement in Medicine: The Analysis of Method Comparison Studies. *Journal of the Royal Statistical Society Series D (The Statistician)*. 1983;32(3):307-17.
55. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet*. 1986;1(8476):307-10.
56. Giavarina D. Understanding Bland Altman analysis. *Biochem Med (Zagreb)*. 2015;25(2):141-51.
57. Shek DT, Ma CM. Longitudinal data analyses using linear mixed models in SPSS: concepts, procedures and illustrations. *ScientificWorldJournal*. 2011;11:42-76.
58. Hales PW, Kirkham FJ, Clark CA. A general model to calculate the spin-lattice (T1) relaxation time of blood, accounting for haematocrit, oxygen saturation and magnetic field strength. *J Cereb Blood Flow Metab*. 2016;36(2):370-4.
59. Wang MS, Haynor DR, Wilson GJ, Maki JH. Intravascular hematocrit layering in equilibrium phase contrast-enhanced MR angiography of the peripheral vasculature. *J Magn Reson Imaging*. 2006;24(6):1393-400.
60. Kanda T, Ishii K, Kawaguchi H, Kitajima K, Takenaka D. High signal intensity in the dentate nucleus and globus pallidus on unenhanced T1-weighted MR images: relationship with increasing cumulative dose of a gadolinium-based contrast material. *Radiology*. 2014;270(3):834-41.

10. Papers I – III

Effect of Drinking Oxygenated Water Assessed by in vivo MRI Relaxometry

Svein Are Sirirud Vatnehol, MSc,^{1,2,3*} Per Kristian Hol, PhD,^{1,2} Atle Bjørnerud, PhD,^{4,5}
Mahmood Amiry-Moghaddam, PhD,^{1,3} Camilla Haglerød, PhD,³ and
Trygve Holck Storås, PhD⁶

Grant Support: This project was funded by the Research Council of Norway.

Background: Oxygen uptake through the gastrointestinal tract after oral administration of oxygenated water in humans is not well studied and is debated in the literature. Due to the paramagnetic properties of oxygen and deoxyhemoglobin, MRI as a technique might be able to detect changes in relaxometry values caused by increased oxygen levels in the blood.

Purpose: To assess whether oxygen dissolved in water is absorbed from the gastrointestinal tract and transported into the bloodstream after oral administration.

Study Type: A randomized, double-blinded, placebo-controlled crossover trial.

Population/Subjects: Thirty healthy male volunteers age 20–35.

Field Strength/Sequence: 3T/Modified Look–Locker inversion recovery (MOLLI) T₁-mapping and multi fast field echo (mFFE) T₂*-mapping.

Assessment: Each volunteer was scanned in two separate sessions. T₁ and T₂* maps were acquired repeatedly covering the hepatic portal vein (HPV) and vena cava inferior (VCI, control vein) before and after intake of oxygenated or control water. Assessments were done by placing a region of interest in the HPV and VCI.

Statistical Test: A mixed linear model was performed to the compare control vs. oxygen group.

Results: Drinking caused a mean 1.6% 95% CI (1.1–2.0% $P < 0.001$) increase in T₁ of HPV blood and water oxygenation attributed another 0.70% 95% confidence interval (CI) (0.07–1.3% $P = 0.028$) increase. Oxygenation did not change T₁ in VCI blood. Mean T₂* increased 9.6% 95% CI (1.7–17.5% $P = 0.017$) after ingestion of oxygenated water and 1.2% 95% CI (–4.3–6.8% $P = 0.661$) after ingestion of control water. The corresponding changes in VCI blood were not significant.

Data Conclusion: Ingestion of water caused changes in T₁ and T₂* of HPV blood compatible with dilution due to water absorption. The effects were enhanced by oxygen. Assessment of oxygen enrichment of HPV blood was not possible due to the dilution effect.

Level of Evidence: 2

Technical Efficacy Stage: 2

J. MAGN. RESON. IMAGING 2020;52:720–728.

Oxygen is vital for the human body and the majority of gas uptake is through the lungs. It has been suggested that oxygen can be absorbed through the intestines, providing an extra source of oxygen delivery.^{1–9} Some groups propose that oxygen-enriched water may play a role in cases such as impaired liver function or at restitution after exercise.^{6,8} Other groups suggest that epithelial cells in the gut would most likely consume the excess oxygen in the gastrointestinal (GI) tract.⁹

Two challenges when evaluating oxygen absorption from water in the GI tract are where to perform measurements and

how to measure a potential uptake. Since the hepatic portal vein (HPV) receives blood from the intestine segments, it may be a suitable vessel for detecting a potential uptake and increase of oxygen in the bloodstream. Similarly, the vena cava inferior (VCI) does not receive blood from the GI system until after the blood has gone through the liver. Therefore, at the level of HPV, we do not expect any changes in oxygen concentrations in VCI and this vessel could function as a control.

In 2001, Forth and Adam⁸ measured oxygen partial pressure (pO₂) changes in the abdominal cavity and the

View this article online at wileyonlinelibrary.com. DOI: 10.1002/jmri.27104

Received Nov 18, 2019, Accepted for publication Feb 12, 2020.

*Address reprint requests to: S. A. S. V. Faculty of medicine, Universit of Oslo, Oslo, Norway. E-mail: svavat@ous-hf.no

From the ¹Faculty of Medicine, University of Oslo, Oslo, Norway; ²The Intervention Centre, Oslo University Hospital, Oslo, Norway; ³Oxy Solutions AS, Oslo, Norway; ⁴Department of Physics, University of Oslo, Oslo, Norway; ⁵Division of Radiology and Nuclear Medicine, Computational Radiology and Artificial Intelligence, Oslo University Hospital, Oslo, Norway; and ⁶Department of Diagnostic Physics, Oslo University Hospital, Oslo, Norway

This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

portal vein of rabbits using direct measurements from probes following oxygenated water administration by gastric tubes. While probes, or blood samples from the gastric veins, may give accurate estimates of O_2 levels, such a procedure is invasive and thus not suitable for human studies; a noninvasive method of evaluation is therefore desirable. In magnetic resonance imaging (MRI), oxygen challenges allow detection and characterization of tissue response to variations in oxygen concentrations.^{10,11} Increased oxygen content in the blood may be achieved by either increasing the dissolved oxygen concentration (DO), increasing the hemoglobin saturation (sO_2), or a combination of the two. The ratio between sO_2 and pO_2/DO is dynamic and governed by the hemoglobin oxygen dissociation curve.^{12,13}

Both DO and deoxyhemoglobin are paramagnetic, while oxyhemoglobin is weakly diamagnetic, as is blood plasma. Since hemoglobin is confined to the intracellular space of the erythrocytes, a conversion from oxyhemoglobin to deoxyhemoglobin creates a heterogeneous susceptibility^{14,15} causing a shortening in T_2^* . Hence, venous blood has a lower T_2^* than arterial blood.

Dissolved oxygen has a shortening effect on the T_1 of plasma. As oxygen diffuses into erythrocytes and is absorbed by deoxyhemoglobin to form oxyhemoglobin, the T_1 relaxation effect ceases and so does the T_2^* relaxation caused by the hemoglobin.^{11,16} These effects could potentially be measured quantitatively by MR relaxometry (MR_r).

The goal of this study was to investigate the effect of drinking oxygen-enriched water by measuring T_1 and T_2^* of the HPV blood draining the GI tract.

Materials and Methods

A randomized, double-blinded, placebo-controlled crossover trial was designed to address the aim of this study. The study was approved by the regional Ethics Committee and all volunteers signed a written informed consent form before inclusion in the study. Thirty male volunteers (aged 20–35) with no regular medications or known liver or blood disease were recruited. Male volunteers were chosen to avoid variation caused by differences in metabolic rates between genders, periodic changes in blood values as a part of the menstrual cycle, as well as avoiding scanning volunteers with unknown pregnancies. All volunteers fasted for a minimum of 6 hours before the examination. The volunteers were randomized with respect to which examination they received: control or oxygenated water. Both volunteers and the investigator were blinded during the whole study. The time between examination days was a minimum of 1 week.

Oxygenated water was produced by a patented process (WO 2016/071691 Oxy Solutions, Oslo, Norway). The same source of water was used for both oxygenated water and control (nonoxygenated) water. Before each examination, water bottles were placed in a heating cabinet to reach a temperature of 37°C . This was performed in order to minimize temperature effects, which are known to alter relaxation times, most notably the effect on T_1 , which increases with temperature in the physiological range.¹⁷

TABLE 1. Selected Sequence Parameters for the MOLLI 10(5)5 Sequence and mFFE Sequence

Sequence	Selected sequence parameters									
	TR (msec)	TE (msec)	Flip angle	Min T_1 (msec)	Max T_1 (msec)	Min TE (msec)	Max TE (msec)	# Echoes	FOV (mm)	Acquisition resolution (mm)
MOLLI 10(5)5	2.4	1.07	20°	187	9187	NA	NA	NA	360 $\times 360$	$1.5 \times 1.5 \times 10$
mFFE	70	NA	25°	NA	NA	1.22	69.22	18	360×360	$1.56 \times 1.55 \times 10$
										Reconstruction resolution (mm)
										0.75 $\times$ 0.75 $\times$ 10
										Scan Duration (s)
										20
										0.94 $\times$ 0.94 $\times$ 10
										17

MRI

A Philips Ingenia 3T scanner (Philips Medical Systems, Best, The Netherlands) was used for MRI experiments.

A modified Look–Locker inversion recovery (MOLLI) with a 10(5)5 scheme and a multi fast field echo (mFFE) was used to acquire T_1 - and T_2^* -weighted images used for T_1 and T_2^* relaxation time measurements. Selected sequence parameters for the sequences are given in Table 1. The MOLLI sequence used a simulated ECG signal set to 60 bpm for triggering and had a breath-hold duration of 20 seconds. The mFFE sequence used no triggering and had a breath-hold duration of 17 seconds.

Following localizer images, two image planes were identified: 1) at a right angle to the HPV (Fig. 1), and 2) at a right angle to the VCI (Fig. 2). Following baseline scans, consisting of eight repetitions of MOLLI and mFFE in each plane, the volunteers drank two bottles (~700 mL in total) of either control or oxygenated water. The volunteers were allowed to sit up for the drinking process but were instructed not to move their buttocks. The volunteers drank the test fluid at their own pace, as rapidly as comfortable. After the volunteers

completed drinking they were assisted to lie straight down. Repeatable slice positioning was confirmed by repeating two of the localizer scans and visually confirming slice placement. In case of motion, the slices were replanned. The time from finished drinking to the first postdrinking scan was noted, and volunteers were scanned for ~30 minutes after the first postdrink scan.

Confirmation of Oxygen Levels

Validation of water type was performed by measuring the T_1 relaxation time in the gastric ventricle at the first timepoint after drinking. After all the experiments were concluded, DO levels were measured in the oxygen bottles from the same batch by using Winkler titration, as described elsewhere.¹⁸

Image Analysis

Parametric T_1 and T_2^* maps were generated in nordicICE (NordicNeuroLab, Bergen, Norway) using a Levenberg–Marquardt nonlinear least-squares algorithm¹⁹ applied to the appropriate expressions for the respective relaxation rates. In each

FIGURE 1: Representative images showing the slice planning of orientation «A». **a,b:** Coronal and orthogonal localizer showing the slice (red line). **c:** A representative slice from the MOLLI 10(5)5 sequence. **D:** The corresponding R_1 map with a representative ROI (white circle) placed in the hepatic portal vein. Zoomed region of "d" is shown as a red box in "c."

FIGURE 2: Representative images showing the slice planning of orientation «B». **a,b:** Coronal and sagittal localizer showing the slice (red line). **c:** Representative slice from the MOLLI 10(5)5 sequence. **d:** The corresponding R_1 map with representative ROI's («V» Vena Cava Inferior and «P» Portal Vein). Zoomed region of "d" is shown as a red box in "c."

parametric map, three regions of interest (ROIs) were placed in the HPV and VCI in their respective planes. Care was taken to include as much of the vessel as possible while avoiding vessel walls and areas of visual artifacts.

Statistical Methods

Following image analysis, average T_1 and T_2^* relaxation times for each blood vessel in each parametric map were calculated. To compensate for outliers in the baseline, the first and third quartile (Q1, Q3) and the interquartile range (IQR) of each session were calculated. Outliers were defined as values outside $Q1 - (2 \times IQR)$ and $Q3 + 2 \times (IQR)$ and were excluded from further analysis. Since there is a normal variation in baseline T_1 and T_2^* relaxation time between individuals, normalized values were calculated by dividing the measured T_1 and T_2^* by the baseline values.

Mixed linear model (MLM) was performed in SPSS v. 25 (IBM, release 2017; Statistics for Windows, Armonk, NY) to compare average differences between drinking oxygenated (oxygen group) and control water (control group). In the MLM model, the normalized T_1 and T_2^* values were set as the dependent variable with

the repeated measures set as the time between measurement and drinking. Fixed factors in the model were the water type, if the measurements were performed pre- or postdrink, and the nested term of water type* pre/postdrink. The repeated covariance type was set to AR1.

Results were considered significant if the P -value was less than 0.05. In addition to the MLM analysis, signal plots of oxygen and control groups were visually evaluated.

Three of the authors are affiliated with a commercial entity that supplied the water samples used in this study. This may constitute a potential conflict of interest. The study presented is part of a collaboration between the commercial entity, the national research council, and a local university. This is a common method of facilitating collaboration between industry and academia in Norway. To ensure that any potential conflict of interest was minimized, all researchers and volunteers were blinded until all analyses were completed.

Results

A total of 133 outliers (6.93%) of the baseline measurements were identified as outliers and removed from the study.

TABLE 2. Output Tables of the Statistical Results From the MLM Analysis

Type III tests of fixed effects ^a					
Sequence		Numerator df	Denominator df	F	Sig.
mFFE portal vein	Intercept	1	1202	10556.857	0.000
	Water	1	1202	5.732	0.017
	Time	1	1202	9.087	0.003
	Time* Water	1	1202	5.732	0.017
mFFE vena cava	Intercept	1	213.365	9816.457	0.000
	Water	1	213.365	0.616	0.433
	Time	1	575.913	4.540	0.034
	Time* Water	1	575.913	1.616	0.204
MOLLI portal vein	Intercept	1	1166.714	1620172.495	0.000
	Water	1	1166.714	4.370	0.037
	Time	1	1168.255	146.960	0.000
	Time* Water	1	1168.255	4.824	0.028
MOLLI vena cava	Intercept	1	182.370	465683.199	0.000
	Water	1	182.370	0.538	0.464
	Time	1	480.607	12.940	0.000
	Time* Water	1	480.607	0.313	0.576

^aDependent variable: Normalised T₁ and T₂* values.
Presenting the type results of the type III test of fixed effects. The table shows that there is a significant effect of drinking water, as well as a significant difference between the oxygen and control group in HPV, but not in VCI.

Average (and mean) T₂* and T₁ values before normalization in the HPV and VCI were 46.53 msec (14.46 msec), 37.98 msec (10.88 msec), 1814 msec (79.5 msec), and 1771 msec (89.7 msec). Tables 2 and 3 present the results of the MLM analysis. Both ingestion and oxygenation show nonzero regression parameters for T₁ of the HPV blood.

The drinking per se was associated with a mean 1.6% 95% CI (1.1–2.0%, $P < 0.001$) increase in T₁ of HPV blood and water oxygenation attributed another 0.70% 95% CI (0.07–1.3%, $P = 0.028$) increase. Correspondingly, drinking was associated with a 0.74% 95% CI (0.07–1.4%, $P = 0.03$) decrease in T₁ of VCI blood, while drinking oxygenated water showed no significant additional effect -0.27% 95% CI (-1.2 – 0.7% , $P = 0.576$). For T₂*, the effect of oxygenated water was an increase in mean value of 9.6% 95% CI (1.7–17.5%, $P = 0.017$) in HPV blood while the change -0.3% 95% CI (-1.2 – 0.7% , $P = 0.204$) in VCI blood was not significant. Drinking per se did not change the mean T₂* of blood in either of the veins (HPV: 1.2% 95% CI (-4.3 – 6.8% , $P = 0.661$) and VCI: 1.4% 95% CI (-3.0 – 5.9% , $P = 0.529$).

Plots showing the normalized T₁ and T₂* values can be found in Fig. 3a–d. It is evident from the plots that the measured effects of oxygenation and drinking are time-dependent. T₁ of HPV blood shows an initial decrease following the ingestion of control water. This decrease was not observed after the ingestion of oxygenated water. Measured T₁ then increases steadily until a peak increase of about 3%. Oxygenated water induces an earlier increase of T₁ and to a higher level than does the control water. T₁ of VCI blood shows an immediate decrease after drinking, and then gradually returns to its baseline value within the first six measurements. There is no apparent effect of water oxygenation on the T₁ of VCI blood.

By visually observing the signal curves: T₂* of the HPV blood shows an increase immediately after the ingestion of water. The increase is higher and the following decrease is slower after the ingestion of oxygenated water. There are no apparent changes in T₂* of VCI blood after ingestion of either water type.

The control measurements (Fig. 4) showed a significant difference in the T₁ measured in the gastric ventricle after

TABLE 3. Edited Output Tables of the Statistical Results From the MLM Analysis

Estimates of fixed effects ^a						
Sequence (measurement)	Fixed effect	Estimate	Std. error	Sig.	95% Confidence interval	
					Lower bound	Upper bound
mFFE portal vein (normalized T_2^* in the hepatic portal vein)	Intercept	1	0.022344	0.000	0.956162	1.043838
	Effect of oxygenated water	0.000%	3.157%	1.000	-6.193%	6.193%
	Effect of drinking	1.244%	2.836%	0.661	-4.320%	6.808%
	Effect of oxygenated water after drinking	9.603%	4.011%	0.017	1.734%	17.471%
mFFE vena cava (normalized T_2^* in the vena cava inferior)	Intercept	0.991188	0.019829	0.000	0.952194	1.030182
	Effect of oxygenated water	-0.518%	2.848%	0.856	-6.119%	5.083%
	Effect of drinking	1.428%	2.269%	0.529	-3.026%	5.883%
	Effect of oxygenated water after drinking	4.224%	3.323%	0.204	-2.303%	10.751%
MOLLI portal vein (normalized T_1 in the hepatic portal vein)	Intercept	1.000056	0.001778	0.000	0.996568	1.003543
	Effect of oxygenated water	-0.017%	0.251%	0.947	-0.509%	0.476%
	Effect of drinking	1.575%	0.225%	0.000	1.133%	2.016%
	Effect of oxygenated water after drinking	0.697%	0.317%	0.028	0.074%	1.319%
MOLLI vena cava (normalized T_1 in the vena cava inferior)	Intercept	1.004278	0.00293	0.000	0.998513	1.010042
	Effect of oxygenated water	-0.079%	0.411%	0.848	-0.889%	0.731%
	Effect of drinking	-0.738%	0.339%	0.030	-1.404%	-0.072%
	Effect of oxygenated water after drinking	-0.272%	0.486%	0.576	-1.226%	0.683%

^aDependent variable: Normalised T_1 and T_2^* values.

Presenting the estimates of the fixed effects, redundant parameters are removed from the table. The table shows that there is a significant effect of drinking water, as well as a significant difference between the oxygen and control group in HPV, but not in VCI.

drinking (oxygenated water: 2596 msec vs. control water: 4217 msec $P < 0.001$, independent samples t -test). Measured DO in the oxygen bottles were 69.34 mg/L (standard deviation 1.67 mg/L) and 7.0 mg/L (standard deviation 0.66 mg/L in the control bottles) the measurements were performed with Winkler titration and after all MRI examinations were completed.

Discussion

In this study we found that the ingestion of water leads to measurable increases in both T_1 and T_2^* relaxation times of HPV blood and this effect is increased with oxygenated water. The differences in relaxation times after intake of oxygenated and control water can be attributed to either of two effects: an increase in oxygen saturation, which would explain

FIGURE 3: Plots of the normalized T_2^* -relaxation times in the hepatic portal vein (HPV) (a), vena cava inferior (VIC) (b), the T_1 relaxation times in the HPV (c), and VCI (d) plotted against time relative to drink. Baseline measurements are grouped as B1-B8 regardless of the time between baseline scans and drinking. The plots show increased values in the hepatic portal vein in the oxygen (gray) compared to the control groups (black). Immediately after drinking there is an increase in the T_2^* -relaxation time in the HPV; this increase is higher when drinking oxygenated water, and persists longer than when drinking control water. For the T_2^* relaxation times in the VCI (b) no such effect is evident. While oxygenated water appears to cause higher relaxation times, this effect was not significant. In "c" we observe an immediate decrease in VCI T_1 -relaxation times following drinking of control water, followed by an increase after the first 6–8 minutes. When drinking oxygenated water, this reduction in T_1 -relaxation times is not evident, and the relaxation times increase faster, and to a higher plateau than when drinking control water. In VCI there is an immediate reduction in T_1 -relaxation times immediately after drinking, followed by a steady increase to baseline; there were no differences between drinking oxygen and control water.

the increase in T_2^* , or that the presence of oxygen increases water absorption, and thus an increase in dilution compared to control water, which could potentially explain the increase in both T_1 and T_2^* relaxation times. As the effect of dilution and oxygenation are opposite with regard to T_1 and equal with regard to T_2^* , we cannot exclude the possibility that some oxygen reached the HPV blood; however, separation of these effects was not attempted in this study.

FIGURE 4: Boxplots showing the average T_1 in the ventricle immediately after drinking.

One alternative explanation of the increase in T_2^* following the intake of oxygenated water is a conversion of the paramagnetic deoxyhemoglobin to the diamagnetic oxyhemoglobin. It may also, however, be a result of susceptibility matching of the blood. Susceptibility matching can occur if a paramagnetic species such as DO is added to the plasma fraction of deoxygenated blood—leading to a reduction in the susceptibility difference (and hence increased T_2^*) between the intracellular and extracellular compartments of whole blood.²⁰ A previous study in rabbits⁸ supports this interpretation, as it showed an increase in pO_2 in the abdominal cavity and HPV after oral administration of oxygenated water. Differences in their results from the current study may be due to several factors, including differences in sample size and methodology in addition to obvious differences in size and metabolism between humans and rabbits. In humans, Nestle et al²¹ previously used conventional T_1 -weighted images to assess the persistence of oxygenation after ingestion of oxygen-saturated water. They concluded that a considerable increase in oxygen concentration could be achieved in the lumen of the stomach. They were, however, not able to follow the oxygen through the digestive tract due to technical limitations.

The increased T_1 time measured in the oxygen group, however, is not compatible with the interpretation that the excess oxygen is transferred to the bloodstream. From the T_1 results, it seems likely that the high O_2 content of the ingested water increases the absorption rate of water and thereby the dilution of HPV blood. While water absorption in the intestines is a passive process, it is dependent on the active transport of ions. The predominant method of transportation of ions is the Na^+/K^+ -ATPase pump, which is dependent on available O_2 .²² Since it is known that reduced levels of O_2 result in a reduction of ATP synthase, it is plausible that increased O_2 levels will result in increased water absorption due to higher activity in the Na^+/K^+ -ATPase pump. These mechanisms, however, need to be studied further to be fully understood. An experiment measuring the time course of intraluminal water volume of the ventricle (and possibly small bowel) could potentially be constructed in the future to find support for this interpretation.

To serve as a control, T_1 and T_2^* in VCI blood were also measured, since this would presumably not be affected by the amount of oxygen ingested. Using the MLM method, we did not observe any effect of water oxygenation on VCI blood. Visually examining the detailed time course after drinking we see that T_1 was significantly reduced for both water qualities. We are not aware of any physiological mechanisms that would explain such a response to water ingestion. This is likely due to the practical act of drinking. Before drinking, the volunteers lay in a stationary position. Slow-flowing blood in the large veins may then partially separate a phenomenon known as intravascular hematocrit (HCT) layering.²³ There is evidence that the plasma layer exhibits a faster flow than the HCT, and downstream venous blood may thereby be diluted and T_1 correspondingly increased. As the volunteers sit up in order to drink, venous blood is presumably mixed and T_1 is decreased in VCI blood. The dilution gradually returns after the volunteer lies down again to continue scanning. The fact that this effect was not seen in the baseline may be due to the duration of volunteer preparation after lying down.

One potential confounder to the in vivo relaxometry measurements performed in this study is the presence of water with varying DO in the ventricle, causing some changes in the local magnetic field. This could potentially lead to systematic differences in the resulting relaxometry measurements between the two groups. Due to the close proximity between HPV and VCI, and the almost equivalent flow rate, we believed that VCI was a sensible vessel to use as a control in this study. Since there was no difference in the changes of relaxometry values caused by drinking oxygenated and control water in the VCI, it is likely that the differences observed in the HPV are a result of changes in the magnetic properties of the blood, and not gross changes in local fields due to the oxygen content of the stomach.

The basic mechanisms of absorption and transportation of water and oxygen were not elucidated in this study, and further studies are required. Our results suggest that the water absorption rate may be influenced by oxygen concentration. Whether or not this effect extends to absorption of other nutrients remains to be investigated. It appears, however, evident that the excess oxygen supplemented in water changes the magnetic properties of HPV blood, and that this change is detectable by MRI relaxometry. Whether these changes are the result of increased intraluminal oxygen content, increased absorption of water, or other factors remains to be seen.

Several other factors apart from oxygen may result in relaxation time changes in the blood. Among these are other solubles, temperature, and blood flow. The experiment was designed with paired examinations and baseline normalization to minimize the effect of most of the confounding factors.

The time of day the volunteers were scanned was not standardized between the examinations, thus there is a possibility that natural fluctuation in baseline saturation levels might occur. The water in both the oxygen and control groups was from the same source, minimizing the possibility of differences in concentration of other solubles. The water samples were also heated to a temperature of 37 °C, which should give less temperature effect than if the water was administered at chilled or room temperature. In terms of flow, we did not measure flow in the HPV during this experiment; further studies would benefit from measuring the flow during the experiments. While it is known that HPV flow is increased after eating or drinking, we believed that this effect should be equal for both groups. The apparently increased dilution connected to oxygenation questions this assumption; however, any effects from flow are presumably still small given the robustness of the applied MRI sequences adapted from cardiac protocols. The effect of flow on T_1 measurements in the portal vein has been investigated earlier²⁴ and it was found that the flow is fast enough to completely replace saturated blood between consecutive readouts. Any further dependence on flow velocity was not investigated.

Limitations

Oxygenated water in the gastric lumen causes changes in the magnetic field. This may have affected the T_1 -measurements.²⁵ Field mapping could have been performed to monitor this effect; however, in the interest of efficient data sampling, in this study this was not performed. Also, with the MOLLI sequence the distance to dark band artifacts gave a visual indication of whether the measurements were off-resonance or not. However, further studies might benefit from including field mapping in the study protocol.

Conclusion

We were able to detect a difference in the T_1 and T_2^* relaxation times in the HPV blood between volunteers consuming either oxygenated or control water. The changes observed were compatible with oxygen causing an increased

rate of water absorption and thereby a higher dilution of HPV blood. Further studies are needed to confirm the observed effect and to address potential applications.

Acknowledgment

The authors thank Jorunn Fraser-Green for practical assistance during the MRI examinations.

References

- Leibetseder V, Strauss-Blasche G, Marktl W, Ekmekcioglu C. Does oxygenated water support aerobic performance and lactate kinetics? *Int J Sports Med* 2006;27(03):232-235.
- McNaughton LR, Kenney S, Siegler J, Midgley AW, Lovell RJ, Bentley DJ. The effect of superoxygenated water on blood gases, lactate, and aerobic cycling performance. *Int J Sports Physiol Perform* 2007;2(4):377-385.
- Piantadosi CA. "Oxygenated" water and athletic performance. *Br J Sports Med* 2006;40(9):740-741.
- Wing-Gaia SL, Subudhi AW, Askew EW. Effects of purified oxygenated water on exercise performance during acute hypoxic exposure. *Int J Sport Nutr Exerc Metab* 2005;15(6):680-688.
- Hampson NB, Pollock NW, Piantadosi CA. Oxygenated water and athletic performance. *JAMA* 2003;290(18):2408-2409.
- Fleming N, Vaughan J, Feeback M. Ingestion of oxygenated water enhances lactate clearance kinetics in trained runners. *J Int Soc Sports Nutr* 2017;14:9.
- Jenkins A, Moreland M, Waddell TB, Fernhall B. Effect of oxygenized water on percent oxygen saturation and performance during exercise. *Med Sci Sports Exerc* 2001;33(5):S167.
- Forth W, Adam O. Uptake of oxygen from the intestine—Experiments with rabbits. *Eur J Med Res* 2001;6(11):488-492.
- Willmert N, Porcari JP, Foster C, Doberstein S, Brice G. The effects of oxygenated water on exercise physiology during incremental exercise and recovery. *J Exerc Physiol Online* 2002;5(4):6.
- Dani KA, Moreton FC, Santosh C, et al. Oxygen challenge magnetic resonance imaging in healthy human volunteers. *J Cereb Blood Flow Metab* 2017;37(1):366-376.
- O'Connor JPB, Robinson SP, Waterton JC. Imaging tumour hypoxia with oxygen-enhanced MRI and BOLD MRI. *Br J Radiol* 2019;92(1096):20180642.
- Collins JA, Rudenski A, Gibson J, Howard L, O'Driscoll R. Relating oxygen partial pressure, saturation and content: The haemoglobin-oxygen dissociation curve. *Breathe (Sheff)* 2015;11(3):194-201.
- Dunn J-O, Mythen M, Grocott M. Physiology of oxygen transport. *BJA Educ* 2016;16(10):341-348.
- Silvennoinen MJ, Clingman CS, Golay X, Kauppinen RA, van Zijl PC. Comparison of the dependence of blood R2 and R2* on oxygen saturation at 1.5 and 4.7 Tesla. *Magn Reson Med* 2003;49(1):47-60.
- Thulborn KR, Waterton JC, Matthews PM, Radda GK. Oxygenation dependence of the transverse relaxation time of water protons in whole blood at high field. *Biochim Biophys Acta* 1982;714(2):265-270.
- Uematsu H, Takahashi M, Hatabu H, et al. Changes in T1 and T2 observed in brain magnetic resonance imaging with delivery of high concentrations of oxygen. *J Comput Assist Tomogr* 2007;31(5):662-665.
- Petrén-Mallmin M, Ericsson A, Rauschnig W, Hemmingsson A. The effect of temperature on MR relaxation times and signal intensities for human tissues. *Magma* 1993;1(3):176-184.
- Vatnehol SAS, Hol PK, Bjørnerud A, Amiry-Moghaddam M, Haglerød C, Storås TH. Determination of oxygen r1 at 3 Tesla using samples with a concentration range of dissolved oxygen. *Magma* 2019 [epub ahead of print]. <https://link.springer.com/article/10.1007/s10334-019-00783-x>
- Press WH, Teukolsky SA, Vetterling WT, Flannery BP. *Numerical recipes in C (2nd ed.): The art of scientific computing*. Cambridge, UK: Cambridge University Press; 1992.
- Bjørnerud A, Briley-Saebo K, Johansson LO, Kellar KE. Effect of NC100150 injection on the (1)H NMR linewidth of human whole blood ex vivo: Dependency on blood oxygen tension. *Magn Reson Med* 2000;44(5):803-807.
- Nestle N, Wunderlich A, Nussle-Kugele K. In vivo observation of oxygen-supersaturated water in the human mouth and stomach. *Magn Reson Imaging* 2004;22(4):551-556.
- Ward JB, Keely SJ, Keely SJ. Oxygen in the regulation of intestinal epithelial transport. *J Physiol* 2014;592(12):2473-2489.
- Wang MS, Haynor DR, Wilson GJ, Maki JH. Intravascular hematocrit layering in equilibrium phase contrast-enhanced MR angiography of the peripheral vasculature. *J Magn Reson Imaging* 2006;24(6):1393-1400.
- Vatnehol SAS, Hol PK, Bjørnerud A, Amiry-Moghaddam M, Haglerød C, Storås TH. Precision of T1-relaxation time measurements in the hepatic portal vein: Influence of measurement technique and sequence parameters. *Magn Reson Mater Phys Biol Med* 2019;32:359-368.
- Kellman P, Hansen MS. T1-mapping in the heart: Accuracy and precision. *J Cardiovasc Magn Reson* 2014;16(1):2.